



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032824\
Data File : VX040818.D
Acq On : 28 Mar 2024 08:57
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Title : SW846 8260
Last Update : Fri Mar 22 02:44:48 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	3297	PASS
75	95	30	60	58.3	8632	PASS
95	95	100	100	100.0	14796	PASS
96	95	5	9	5.7	844	PASS
173	174	0.00	2	1.2	128	PASS
174	95	50	100	74.1	10970	PASS
175	174	5	9	8.3	907	PASS
176	174	95	101	96.1	10543	PASS
177	176	5	9	7.5	795	PASS