

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032825\
 Data File : VX045506.D
 Acq On : 28 Mar 2025 15:18
 Operator : JC/MD
 Sample : VX0328WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0328WBL01

Quant Time: Mar 31 10:22:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML032825WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 28 14:43:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.751	114	150584	50.00	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	134411	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	55732	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	54273	48.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.40%
7) Chloroethane-d5	1.648	69	39336	53.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	106.66%
11) 1,1-Dichloroethene-d2	2.294	65	29617	50.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.96%
21) 2-Butanone-d5	4.452	46	89849	119.94	ug/L	-0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	119.94%
24) Chloroform-d	5.050	84	122434	49.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.00%
26) 1,2-Dichloroethane-d4	5.946	65	88660	50.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.42%
32) Benzene-d6	5.964	84	229095	52.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.14%
36) 1,2-Dichloropropane-d6	7.306	67	74254	52.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.84%
41) Toluene-d8	8.647	98	197878	50.29	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.58%
43) trans-1,3-Dichloroprop...	8.945	79	28601	49.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.28%
47) 2-Hexanone-d5	9.384	63	60660	113.34	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	113.34%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	99520	53.43	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.86%
66) 1,2-Dichlorobenzene-d4	12.317	152	65040	54.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.80%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032825\
Data File : VX045506.D
Acq On : 28 Mar 2025 15:18
Operator : JC/MD
Sample : VX0328WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0328WBL01

Quant Time: Mar 31 10:22:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML032825WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 28 14:43:26 2025
Response via : Initial Calibration

