

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032919\
 Data File : VX008541.D
 Acq On : 29 Mar 2019 14:16
 Operator : JC/SP
 Sample : K2115-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 QNWP8-EQUIP.BLANK

Quant Time: Mar 30 00:11:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	207605	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	348166	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	292641	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	122123	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	145796	46.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.64%	
35) Dibromofluoromethane	5.49	113	106559	46.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.88%	
50) Toluene-d8	8.71	98	430287	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
62) 4-Bromofluorobenzene	11.13	95	142547	42.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.76%	
Target Compounds						
16) Acetone	2.44	43	15858	9.640	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032919\
Data File : VX008541.D
Acq On : 29 Mar 2019 14:16
Operator : JC/SP
Sample : K2115-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
QNWP8-EQUIP.BLANK

Quant Time: Mar 30 00:11:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

