

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032919\
 Data File : VX008536.D
 Acq On : 29 Mar 2019 10:37
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 29 23:52:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.589	0.565	4.1	104	0.00
3 P	Chloromethane	0.841	0.734	12.7	97	0.00
4 C	Vinyl Chloride	0.823	0.770	6.4#	99	0.00
5 T	Bromomethane	0.492	0.367	25.4#	89	0.00
6 T	Chloroethane	0.487	0.468	3.9	97	0.00
7 T	Trichlorofluoromethane	0.932	0.897	3.8	101	0.00
8 T	Diethyl Ether	0.441	0.423	4.1	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.563	0.584	-3.7	109	0.00
10 T	Methyl Iodide	0.505	0.488	3.4	92	0.00
11 T	Tert butyl alcohol	0.168	0.171	-1.8	100	0.00
12 CM	1,1-Dichloroethene	0.582	0.562	3.4#	101	0.00
13 T	Acrolein	0.122	0.152	-24.6#	133	0.00
14 T	Allyl chloride	1.282	1.328	-3.6	107	0.00
15 T	Acrylonitrile	0.440	0.441	-0.2	104	0.00
16 T	Acetone	0.396	0.419	-5.8	110	0.00
17 T	Carbon Disulfide	1.930	1.695	12.2	101	0.00
18 T	Methyl Acetate	0.952	0.958	-0.6	106	0.00
19 T	Methyl tert-butyl Ether	2.096	2.213	-5.6	108	0.00
20 T	Methylene Chloride	0.713	0.697	2.2	106	0.00
21 T	trans-1,2-Dichloroethene	0.637	0.621	2.5	106	0.00
22 T	Diisopropyl ether	2.523	2.637	-4.5	108	0.00
23 T	Vinyl Acetate	2.202	2.322	-5.4	106	0.00
24 P	1,1-Dichloroethane	1.274	1.304	-2.4	106	0.00
25 T	2-Butanone	0.635	0.664	-4.6	107	0.00
26 T	2,2-Dichloropropane	0.847	0.960	-13.3	114	0.00
27 T	cis-1,2-Dichloroethene	0.711	0.736	-3.5	108	0.00
28 T	Bromochloromethane	0.629	0.640	-1.7	108	0.00
29 T	Tetrahydrofuran	0.419	0.417	0.5	104	0.00
30 C	Chloroform	1.150	1.188	-3.3#	107	0.00
31 T	Cyclohexane	1.215	1.270	-4.5	110	0.00
32 T	1,1,1-Trichloroethane	0.919	0.973	-5.9	107	0.00
33 S	1,2-Dichloroethane-d4	0.758	0.785	-3.6	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.326	0.352	-8.0	105	0.00
36 T	1,1-Dichloropropene	0.533	0.568	-6.6	108	0.00
37 T	Ethyl Acetate	0.691	0.736	-6.5	104	0.00
38 T	Carbon Tetrachloride	0.445	0.490	-10.1	107	0.00
39 T	Methylcyclohexane	0.637	0.726	-14.0	113	0.00
40 TM	Benzene	1.596	1.703	-6.7	106	0.00
41 T	Methacrylonitrile	0.388	0.415	-7.0	104	0.00
42 TM	1,2-Dichloroethane	0.584	0.607	-3.9	106	0.00
43 T	Isopropyl Acetate	1.105	1.209	-9.4	107	0.00
44 TM	Trichloroethene	0.375	0.399	-6.4	109	0.00
45 C	1,2-Dichloropropane	0.454	0.493	-8.6#	107	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX032919\
 Data File : VX008536.D
 Acq On : 29 Mar 2019 10:37
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 29 23:52:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.265	0.279	-5.3	107	0.00
47 T	Bromodichloromethane	0.507	0.567	-11.8	108	0.00
48 T	Methyl methacrylate	0.553	0.612	-10.7	107	0.00
49 T	1,4-Dioxane	0.011	0.011	0.0	97	0.00
50 S	Toluene-d8	1.304	1.440	-10.4	106	0.00
51 T	4-Methyl-2-Pentanone	0.717	0.790	-10.2	106	0.00
52 CM	Toluene	0.954	1.038	-8.8#	108	0.00
53 T	t-1,3-Dichloropropene	0.555	0.654	-17.8	110	0.00
54 T	cis-1,3-Dichloropropene	0.632	0.721	-14.1	109	0.00
55 T	1,1,2-Trichloroethane	0.388	0.421	-8.5	107	0.00
56 T	Ethyl methacrylate	0.701	0.789	-12.6	107	0.00
57 T	1,3-Dichloropropane	0.696	0.743	-6.8	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.362	0.414	-14.4	109	0.00
59 T	2-Hexanone	0.556	0.607	-9.2	103	0.00
60 T	Dibromochloromethane	0.369	0.419	-13.6	108	0.00
61 T	1,2-Dibromoethane	0.396	0.430	-8.6	106	0.00
62 S	4-Bromofluorobenzene	0.483	0.543	-12.4	107	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.338	0.383	-13.3	115	0.00
65 PM	Chlorobenzene	1.087	1.205	-10.9	111	0.00
66 T	1,1,1,2-Tetrachloroethane	0.375	0.423	-12.8	108	0.00
67 C	Ethyl Benzene	2.052	2.270	-10.6#	109	0.00
68 T	m/p-Xylenes	0.747	0.831	-11.2	110	0.00
69 T	o-Xylene	0.724	0.803	-10.9	109	0.00
70 T	Styrene	1.251	1.427	-14.1	110	0.00
71 P	Bromoform	0.296	0.345	-16.6	108	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	4.068	4.443	-9.2	111	0.00
74 T	N-amyl acetate	2.354	2.529	-7.4	102	0.00
75 P	1,1,2,2-Tetrachloroethane	1.578	1.601	-1.5	103	0.00
76 T	1,2,3-Trichloropropane	1.486	1.386	6.7	91	0.00
77 T	Bromobenzene	0.987	1.054	-6.8	110	0.00
78 T	n-propylbenzene	4.765	5.316	-11.6	112	0.00
79 T	2-Chlorotoluene	2.898	3.126	-7.9	109	0.00
80 T	1,3,5-Trimethylbenzene	3.439	3.788	-10.1	111	0.00
81 T	trans-1,4-Dichloro-2-butene	0.464	0.547	-17.9	108	0.00
82 T	4-Chlorotoluene	3.343	3.621	-8.3	110	0.00
83 T	tert-Butylbenzene	3.318	3.637	-9.6	111	0.00
84 T	1,2,4-Trimethylbenzene	3.478	3.836	-10.3	111	0.00
85 T	sec-Butylbenzene	3.951	4.425	-12.0	112	0.00
86 T	p-Isopropyltoluene	3.486	4.027	-15.5	113	0.00
87 T	1,3-Dichlorobenzene	1.764	1.928	-9.3	113	0.00
88 T	1,4-Dichlorobenzene	1.780	1.928	-8.3	112	0.00
89 T	n-Butylbenzene	3.303	3.896	-18.0	115	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032919\
Data File : VX008536.D
Acq On : 29 Mar 2019 10:37
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 29 23:52:09 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.581	0.672	-15.7	113	0.00
91 T	1,2-Dichlorobenzene	1.766	1.868	-5.8	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.344	0.340	1.2	98	0.00
93 T	1,2,4-Trichlorobenzene	1.189	1.353	-13.8	113	0.00
94 T	Hexachlorobutadiene	0.565	0.662	-17.2	116	0.00
95 T	Naphthalene	4.077	4.351	-6.7	106	0.00
96 T	1,2,3-Trichlorobenzene	1.195	1.315	-10.0	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6