

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021568.D
 Acq On : 29 Mar 2021 11:46
 Operator : JC/MD
 Sample : M1787-11 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-122D2-20210319

Manual Integrations
 APPROVED

MMDadoda
 3/31/2021 9:41:39 AM

Quant Time: Mar 30 03:25:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	61145	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	109828	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	111860	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	48900	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	50839	49.97	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.94%
35) Dibromofluoromethane	5.464	113	39801	51.02	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.04%
50) Toluene-d8	8.694	98	140601	50.58	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.16%
62) 4-Bromofluorobenzene	11.118	95	56434	52.21	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.42%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.364	101	1467	2.33	ug/l	91
12) 1,1-Dichloroethene	2.358	96	549	0.89	ug/l	96
27) cis-1,2-Dichloroethene	4.564	96	530	0.67	ug/l	58
38) Carbon Tetrachloride	5.741	117	474m	0.41	ug/l	
44) Trichloroethene	7.188	130	534012	662.21	ug/l	97
55) 1,1,2-Trichloroethane	9.200	97	215	0.27	ug/l #	80
64) Tetrachloroethene	9.318	164	581	0.78	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021568.D
Acq On : 29 Mar 2021 11:46
Operator : JC/MD
Sample : M1787-11 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-122D2-20210319

Manual Integrations
APPROVED

MMDadoda
3/31/2021 9:41:39 AM

Quant Time: Mar 30 03:25:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

