

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021570.D
 Acq On : 29 Mar 2021 12:32
 Operator : JC/MD
 Sample : M1787-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP03-20210319

Quant Time: Mar 30 03:25:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	59077	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	109827	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	115653	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	54164	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	51931	52.83	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.66%
35) Dibromofluoromethane	5.464	113	39363	50.46	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.92%
50) Toluene-d8	8.694	98	140469	50.53	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.06%
62) 4-Bromofluorobenzene	11.118	95	58645	54.26	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.52%

Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.187	85	394	0.61	ug/l	82
9) 1,1,2-Trichlorotrifluo...	2.364	101	60727	99.87	ug/l	93
12) 1,1-Dichloroethene	2.358	96	772	1.30	ug/l	89
16) Acetone	2.423	43	1266	3.41	ug/l	96
27) cis-1,2-Dichloroethene	4.564	96	2032	2.68	ug/l	98
30) Chloroform	5.170	83	664	0.50	ug/l #	68
38) Carbon Tetrachloride	5.758	117	714	0.62	ug/l #	79
44) Trichloroethene	7.188	130	166003	205.86	ug/l	92
64) Tetrachloroethene	9.318	164	5444	7.03	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021570.D
Acq On : 29 Mar 2021 12:32
Operator : JC/MD
Sample : M1787-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP03-20210319

Quant Time: Mar 30 03:25:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

