

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021577.D
 Acq On : 29 Mar 2021 15:16
 Operator : JC/MD
 Sample : M1787-17DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP03-20210319DL

Quant Time: Mar 30 03:28:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	56661	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	107573	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	110424	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	48953	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	51484	54.61	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.22%	
35) Dibromofluoromethane	5.464	113	39406	51.57	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.14%	
50) Toluene-d8	8.694	98	135920	49.92	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.84%	
62) 4-Bromofluorobenzene	11.118	95	56025	52.92	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.84%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.364	101	13150	22.55	ug/l	97
12) 1,1-Dichloroethene	2.346	96	234	0.41	ug/l #	38
27) cis-1,2-Dichloroethene	4.570	96	559	0.77	ug/l	88
44) Trichloroethene	7.188	130	38956	49.32	ug/l	95
64) Tetrachloroethene	9.318	164	1439	1.95	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021577.D
Acq On : 29 Mar 2021 15:16
Operator : JC/MD
Sample : M1787-17DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP03-20210319DL

Quant Time: Mar 30 03:28:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

