

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021582.D
 Acq On : 29 Mar 2021 17:13
 Operator : JC/MD
 Sample : M1860-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-032621

Quant Time: Mar 30 03:30:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.628	168	63263	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	115027	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	115396	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	48213	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	53027	50.38	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.76%	
35) Dibromofluoromethane	5.470	113	40500	49.57	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.14%	
50) Toluene-d8	8.694	98	146344	50.26	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.52%	
62) 4-Bromofluorobenzene	11.117	95	56700	50.09	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.18%	
Target Compounds						
16) Acetone	2.422	43	3942	9.90	ug/l	# 83
20) Methylene Chloride	2.834	84	1195	1.54	ug/l	# 88
84) 1,2,4-Trimethylbenzene	11.788	105	1227	0.42	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021582.D
Acq On : 29 Mar 2021 17:13
Operator : JC/MD
Sample : M1860-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-032621

Quant Time: Mar 30 03:30:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

