

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021586.D
 Acq On : 29 Mar 2021 18:47
 Operator : JC/MD
 Sample : M1862-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Mar 30 03:31:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	60084	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	114078	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	115963	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	49036	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	53113	53.13	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 106.26%		
35) Dibromofluoromethane	5.464	113	40313	49.75	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.50%		
50) Toluene-d8	8.694	98	144250	49.95	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.90%		
62) 4-Bromofluorobenzene	11.118	95	57133	50.89	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 101.78%		
Target Compounds						
16) Acetone	2.423	43	1464	3.87	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021586.D
Acq On : 29 Mar 2021 18:47
Operator : JC/MD
Sample : M1862-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Mar 30 03:31:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

