

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021588.D
 Acq On : 29 Mar 2021 19:34
 Operator : JC/MD
 Sample : M1862-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Mar 30 03:32:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	55791	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	104044	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	106314	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	48863	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	49099	52.89	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.78%	
35) Dibromofluoromethane	5.464	113	37155	50.27	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.54%	
50) Toluene-d8	8.694	98	131264	49.84	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.68%	
62) 4-Bromofluorobenzene	11.118	95	54341	53.07	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.14%	
Target Compounds						
16) Acetone	2.423	43	1701	4.85	ug/l	# 86
19) Methyl tert-butyl Ether	3.164	73	1785	0.85	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021588.D
Acq On : 29 Mar 2021 19:34
Operator : JC/MD
Sample : M1862-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Mar 30 03:32:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

