

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021596.D
 Acq On : 29 Mar 2021 23:30
 Operator : JC/MD
 Sample : M1843-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-20210325

Quant Time: Mar 30 05:21:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	60035	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	107497	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	108994	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	48184	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	50539	50.59	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.18%	
35) Dibromofluoromethane	5.464	113	38515	50.44	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.88%	
50) Toluene-d8	8.694	98	138603	50.94	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.88%	
62) 4-Bromofluorobenzene	11.118	95	54707	51.71	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.42%	
Target Compounds						
						Qvalue
16) Acetone	2.423	43	6397	16.94	ug/l	# 84
20) Methylene Chloride	2.834	84	1015	1.38	ug/l	# 83
52) Toluene	8.765	92	20503	11.10	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021596.D
Acq On : 29 Mar 2021 23:30
Operator : JC/MD
Sample : M1843-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-20210325

Quant Time: Mar 30 05:21:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

