

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021597.D
 Acq On : 29 Mar 2021 23:53
 Operator : JC/MD
 Sample : M1843-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20210325

Quant Time: Mar 30 05:21:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	60289	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	110522	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	112388	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	47318	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	51630	51.47	ug/l	0.00
Spiked Amount 50.000	Range	78 - 117	Recovery	= 102.94%		
35) Dibromofluoromethane	5.464	113	39944	50.88	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	= 101.76%		
50) Toluene-d8	8.694	98	141303	50.51	ug/l	0.00
Spiked Amount 50.000	Range	92 - 112	Recovery	= 101.02%		
62) 4-Bromofluorobenzene	11.118	95	55840	51.34	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	= 102.68%		
Target Compounds						
16) Acetone	2.417	43	6351	16.74	ug/l #	83
20) Methylene Chloride	2.834	84	968	1.31	ug/l #	86
52) Toluene	8.765	92	18555	9.77	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021597.D
Acq On : 29 Mar 2021 23:53
Operator : JC/MD
Sample : M1843-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20210325

Quant Time: Mar 30 05:21:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

