

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021600.D
 Acq On : 30 Mar 2021 01:03
 Operator : JC/MD
 Sample : M1843-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D2-20210324

Quant Time: Mar 30 05:22:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	60062	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	109758	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	109587	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	47697	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	52255	52.29	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.58%
35) Dibromofluoromethane	5.464	113	39839	51.10	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.20%
50) Toluene-d8	8.694	98	139396	50.17	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.34%
62) 4-Bromofluorobenzene	11.118	95	55964	51.81	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.62%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.358	96	2285	3.77	ug/l	95
16) Acetone	2.417	43	2138	5.66	ug/l #	83
21) trans-1,2-Dichloroethene	3.146	96	402	0.60	ug/l	89
24) 1,1-Dichloroethane	3.670	63	1697	1.32	ug/l #	81
27) cis-1,2-Dichloroethene	4.564	96	40847	52.91	ug/l	98
44) Trichloroethene	7.188	130	340319	422.29	ug/l	98
45) 1,2-Dichloropropane	7.488	63	1292	1.56	ug/l	90
64) Tetrachloroethene	9.318	164	24042	32.75	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021600.D
Acq On : 30 Mar 2021 01:03
Operator : JC/MD
Sample : M1843-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D2-20210324

Quant Time: Mar 30 05:22:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

