

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021603.D
 Acq On : 30 Mar 2021 02:14
 Operator : JC/MD
 Sample : M1843-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP06-20210324

Quant Time: Mar 30 05:23:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	57129	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.829	114	105760	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	113219	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	51674	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	50371	52.99	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	105.98%	
35) Dibromofluoromethane	5.458	113	37934	50.50	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.00%	
50) Toluene-d8	8.694	98	135243	50.52	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	101.04%	
62) 4-Bromofluorobenzene	11.118	95	56182	53.98	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	107.96%	
Target Compounds						Qvalue
44) Trichloroethene	7.182	130	8748	11.27	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021603.D
Acq On : 30 Mar 2021 02:14
Operator : JC/MD
Sample : M1843-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP06-20210324

Quant Time: Mar 30 05:23:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

