

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021605.D
 Acq On : 30 Mar 2021 03:01
 Operator : JC/MD
 Sample : M1843-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE117D2-20210325

Quant Time: Mar 30 05:23:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	60265	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	109855	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	110457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	47086	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	52075	51.93	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.86%	
35) Dibromofluoromethane	5.464	113	39831	51.04	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.08%	
50) Toluene-d8	8.694	98	138406	49.77	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.54%	
62) 4-Bromofluorobenzene	11.118	95	53975	49.93	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.86%	
Target Compounds						
						Qvalue
16) Acetone	2.423	43	1547	4.08	ug/l	89
44) Trichloroethene	7.194	130	988	1.22	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021605.D
Acq On : 30 Mar 2021 03:01
Operator : JC/MD
Sample : M1843-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE117D2-20210325

Quant Time: Mar 30 05:23:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

