

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021608.D
 Acq On : 30 Mar 2021 04:11
 Operator : JC/MD
 Sample : M1843-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D2-20210325

Quant Time: Apr 09 06:53:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	59436	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	110344	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	117448	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	53372	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	52337	52.92	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.84%
35) Dibromofluoromethane	5.464	113	39562	50.47	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.94%
50) Toluene-d8	8.694	98	141158	50.54	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.08%
62) 4-Bromofluorobenzene	11.118	95	59228	54.54	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.08%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.364	101	37923	61.99	ug/l	97
12) 1,1-Dichloroethene	2.352	96	17246	28.79	ug/l	86
24) 1,1-Dichloroethane	3.670	63	4593	3.61	ug/l	99
27) cis-1,2-Dichloroethene	4.570	96	3854	5.04	ug/l	97
30) Chloroform	5.176	83	2326	1.73	ug/l	85
32) 1,1,1-Trichloroethane	5.464	97	2911	2.44	ug/l #	94
38) Carbon Tetrachloride	5.747	117	3806	3.28	ug/l #	89
44) Trichloroethene	7.188	130	733649	905.52	ug/l	95
55) 1,1,2-Trichloroethane	9.200	97	1313	1.64	ug/l	92
64) Tetrachloroethene	9.318	164	414	0.53	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021608.D
Acq On : 30 Mar 2021 04:11
Operator : JC/MD
Sample : M1843-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D2-20210325

Quant Time: Apr 09 06:53:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

