

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021609.D
 Acq On : 30 Mar 2021 04:34
 Operator : JC/MD
 Sample : M1843-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20210325

Quant Time: Mar 30 05:24:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	58275	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	108247	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	111099	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	46834	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	51596	53.21	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.42%	
35) Dibromofluoromethane	5.470	113	39358	51.19	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.38%	
50) Toluene-d8	8.694	98	139264	50.83	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.66%	
62) 4-Bromofluorobenzene	11.118	95	55479	52.08	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.16%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.187	85	957	1.49	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.364	101	18201	30.34	ug/l	96
12) 1,1-Dichloroethene	2.352	96	2966	5.05	ug/l	81
24) 1,1-Dichloroethane	3.670	63	1790	1.43	ug/l #	94
27) cis-1,2-Dichloroethene	4.558	96	3846	5.13	ug/l	99
30) Chloroform	5.182	83	3784	2.87	ug/l	89
32) 1,1,1-Trichloroethane	5.458	97	556	0.48	ug/l #	48
38) Carbon Tetrachloride	5.753	117	2897	2.54	ug/l	98
44) Trichloroethene	7.188	130	366646	461.31	ug/l	93
49) 1,4-Dioxane	7.729	88	325	17.09	ug/l #	70
55) 1,1,2-Trichloroethane	9.194	97	1139	1.45	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021609.D
Acq On : 30 Mar 2021 04:34
Operator : JC/MD
Sample : M1843-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D1-20210325

Quant Time: Mar 30 05:24:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

