

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021610.D
 Acq On : 30 Mar 2021 04:58
 Operator : JC/MD
 Sample : M1843-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D2-20210325

Quant Time: Mar 30 05:25:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	57212	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.823	114	105970	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	113027	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	51139	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	50659	53.22	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.44%	
35) Dibromofluoromethane	5.458	113	38130	50.66	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.32%	
50) Toluene-d8	8.694	98	137395	51.22	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.44%	
62) 4-Bromofluorobenzene	11.118	95	55686	53.40	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.80%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.364	101	10800	18.34	ug/l	96
16) Acetone	2.423	43	1513	4.20	ug/l	92
27) cis-1,2-Dichloroethene	4.564	96	1313	1.79	ug/l	89
30) Chloroform	5.176	83	820	0.63	ug/l	100
44) Trichloroethene	7.188	130	78396	100.76	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021610.D
Acq On : 30 Mar 2021 04:58
Operator : JC/MD
Sample : M1843-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D2-20210325

Quant Time: Mar 30 05:25:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

