

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021615.D
 Acq On : 30 Mar 2021 06:54
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 30 07:16:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	0.550	0.513	6.7	98	0.00
3 P	Chloromethane	0.702	0.606	13.7	105	0.00
4 C	Vinyl Chloride	0.562	0.562	0.0	104	0.00
5 T	Bromomethane	0.226	0.251	-11.1	121	0.00
6 T	Chloroethane	0.357	0.351	1.7	104	0.00
7 T	Trichlorofluoromethane	0.914	0.924	-1.1	104	0.00
8 T	Diethyl Ether	0.317	0.310	2.2	105	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.515	0.498	3.3	103	0.00
10 T	Methyl Iodide	0.362	0.466	-28.7#	136	0.00
11 T	Tert butyl alcohol	0.138	0.136	1.4	108	0.00
12 CM	1,1-Dichloroethene	0.504	0.487	3.4#	107	0.00
13 T	Acrolein	0.077	0.042	45.5#	67	0.00
14 T	Allyl chloride	1.023	0.951	7.0	99	0.00
15 T	Acrylonitrile	0.327	0.325	0.6	104	0.00
16 T	Acetone	0.315	0.314	0.3	108	0.00
17 T	Carbon Disulfide	1.415	1.220	13.8	96	0.00
18 T	Methyl Acetate	0.909	0.986	-8.5	111	0.00
19 T	Methyl tert-butyl Ether	1.882	1.914	-1.7	106	0.00
20 T	Methylene Chloride	0.614	0.581	5.4	103	0.00
21 T	trans-1,2-Dichloroethene	0.553	0.543	1.8	104	0.00
22 T	Diisopropyl ether	1.987	2.045	-2.9	107	0.00
23 T	Vinyl Acetate	1.681	1.686	-0.3	103	0.00
24 P	1,1-Dichloroethane	1.071	1.084	-1.2	106	0.00
25 T	2-Butanone	0.477	0.475	0.4	105	0.00
26 T	2,2-Dichloropropane	0.928	0.570	38.6#	64	0.00
27 T	cis-1,2-Dichloroethene	0.643	0.647	-0.6	106	0.00
28 T	Bromochloromethane	0.536	0.561	-4.7	118	0.00
29 T	Tetrahydrofuran	0.309	0.311	-0.6	107	0.00
30 C	Chloroform	1.131	1.171	-3.5#	108	0.00
31 T	Cyclohexane	0.915	0.886	3.2	100	0.00
32 T	1,1,1-Trichloroethane	1.002	1.029	-2.7	107	0.00
33 S	1,2-Dichloroethane-d4	0.832	0.878	-5.5	119	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	112	0.00
35 S	Dibromofluoromethane	0.355	0.365	-2.8	122	0.00
36 T	1,1-Dichloropropene	0.490	0.459	6.3	104	0.00
37 T	Ethyl Acetate	0.566	0.520	8.1	106	0.00
38 T	Carbon Tetrachloride	0.526	0.519	1.3	107	0.00
39 T	Methylcyclohexane	0.543	0.488	10.1	96	0.00
40 TM	Benzene	1.391	1.358	2.4	106	0.00
41 T	Methacrylonitrile	0.329	0.308	6.4	105	0.00
42 TM	1,2-Dichloroethane	0.603	0.587	2.7	106	0.00
43 T	Isopropyl Acetate	0.920	0.871	5.3	103	0.00
44 TM	Trichloroethene	0.367	0.364	0.8	110	0.00
45 C	1,2-Dichloropropane	0.377	0.368	2.4#	105	0.00
46 T	Dibromomethane	0.275	0.262	4.7	104	0.00
47 T	Bromodichloromethane	0.559	0.554	0.9	107	0.00
48 T	Methyl methacrylate	0.491	0.470	4.3	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021615.D
 Acq On : 30 Mar 2021 06:54
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 30 07:16:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	110	0.00
50 S	Toluene-d8	1.266	1.265	0.1	114	0.00
51 T	4-Methyl-2-Pentanone	0.566	0.564	0.4	106	0.00
52 CM	Toluene	0.859	0.850	1.0#	105	0.00
53 T	t-1,3-Dichloropropene	0.564	0.522	7.4	98	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.555	7.7	98	0.00
55 T	1,1,2-Trichloroethane	0.363	0.363	0.0	108	0.00
56 T	Ethyl methacrylate	0.548	0.547	0.2	103	0.00
57 T	1,3-Dichloropropane	0.621	0.614	1.1	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.307	0.298	2.9	103	0.00
59 T	2-Hexanone	0.425	0.423	0.5	106	0.00
60 T	Dibromochloromethane	0.406	0.422	-3.9	110	0.00
61 T	1,2-Dibromoethane	0.390	0.386	1.0	105	0.00
62 S	4-Bromofluorobenzene	0.492	0.495	-0.6	117	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	110	0.00
64 T	Tetrachloroethene	0.335	0.343	-2.4	110	0.00
65 PM	Chlorobenzene	1.004	0.958	4.6	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.381	1.0	105	0.00
67 C	Ethyl Benzene	1.831	1.788	2.3#	104	0.00
68 T	m/p-Xylenes	0.652	0.653	-0.2	105	0.00
69 T	o-Xylene	0.632	0.632	0.0	106	0.00
70 T	Styrene	1.086	1.078	0.7	105	0.00
71 P	Bromoform	0.313	0.312	0.3	106	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73 T	Isopropylbenzene	3.513	3.639	-3.6	104	0.00
74 T	N-amyl acetate	1.700	1.709	-0.5	106	0.00
75 P	1,1,2,2-Tetrachloroethane	1.241	1.242	-0.1	106	0.00
76 T	1,2,3-Trichloropropane	1.164	1.181	-1.5	103	0.00
77 T	Bromobenzene	0.871	0.893	-2.5	106	0.00
78 T	n-propylbenzene	4.149	4.241	-2.2	104	0.00
79 T	2-Chlorotoluene	2.517	2.571	-2.1	105	0.00
80 T	1,3,5-Trimethylbenzene	3.001	3.121	-4.0	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.382	0.332	13.1	90	0.00
82 T	4-Chlorotoluene	2.988	3.018	-1.0	104	0.00
83 T	tert-Butylbenzene	2.922	2.994	-2.5	105	0.00
84 T	1,2,4-Trimethylbenzene	3.026	3.175	-4.9	106	0.00
85 T	sec-Butylbenzene	3.371	3.469	-2.9	103	0.00
86 T	p-Isopropyltoluene	3.068	3.126	-1.9	100	0.00
87 T	1,3-Dichlorobenzene	1.568	1.598	-1.9	108	0.00
88 T	1,4-Dichlorobenzene	1.622	1.614	0.5	108	0.00
89 T	n-Butylbenzene	2.811	2.751	2.1	98	0.00
90 T	Hexachloroethane	0.554	0.568	-2.5	107	0.00
91 T	1,2-Dichlorobenzene	1.525	1.551	-1.7	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.286	0.286	0.0	102	0.00
93 T	1,2,4-Trichlorobenzene	0.999	1.007	-0.8	106	0.00
94 T	Hexachlorobutadiene	0.451	0.420	6.9	100	0.00
95 T	Naphthalene	3.259	3.377	-3.6	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
Data File : VX021615.D
Acq On : 30 Mar 2021 06:54
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 30 07:16:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.999	1.019	-2.0	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5