

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032922\
 Data File : VX027793.D
 Acq On : 29 Mar 2022 23:03
 Operator : JC/MD
 Sample : N2041-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE117D2-20220317

Quant Time: Mar 30 05:16:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	81859	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	145580	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120561	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51265	48.537	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.080%
35) Dibromofluoromethane	5.391	113	48370	51.475	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.940%
50) Toluene-d8	8.653	98	162340	49.101	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.200%
62) 4-Bromofluorobenzene	11.079	95	62956	44.808	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032922\
Data File : VX027793.D
Acq On : 29 Mar 2022 23:03
Operator : JC/MD
Sample : N2041-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE117D2-20220317

Quant Time: Mar 30 05:16:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

