

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032922\
 Data File : VX027794.D
 Acq On : 29 Mar 2022 23:26
 Operator : JC/MD
 Sample : N2041-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW4-20220317

Quant Time: Mar 30 05:17:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79298	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	143063	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	50193	49.057	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.120%
35) Dibromofluoromethane	5.385	113	47400	51.330	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.660%
50) Toluene-d8	8.647	98	167480	51.547	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.100%
62) 4-Bromofluorobenzene	11.079	95	64898	47.003	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.000%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	4249	5.722	ug/l	98
12) 1,1-Dichloroethene	2.325	96	1385	2.033	ug/l	85
16) Acetone	2.380	43	1221	2.696	ug/l	# 76
27) cis-1,2-Dichloroethene	4.495	96	1521	1.706	ug/l	95
30) Chloroform	5.080	83	698	0.445	ug/l	# 44
38) Carbon Tetrachloride	5.678	117	509	0.396	ug/l	# 64
44) Trichloroethene	7.129	130	765029	798.682	ug/l	98
49) 1,4-Dioxane	7.678	88	350	14.639	ug/l	# 81
55) 1,1,2-Trichloroethane	9.159	97	306	0.311	ug/l	# 86
64) Tetrachloroethene	9.275	164	3915	4.642	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032922\
Data File : VX027794.D
Acq On : 29 Mar 2022 23:26
Operator : JC/MD
Sample : N2041-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW4-20220317

Quant Time: Mar 30 05:17:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

