

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032922\
 Data File : VX027783.D
 Acq On : 29 Mar 2022 18:44
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX032922

Quant Time: Mar 30 04:41:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.453	0.496	-9.5	99	0.00
3 P	Chloromethane	0.400	0.407	-1.7	102	0.00
4 C	Vinyl Chloride	0.411	0.420	-2.2#	100	0.00
5 T	Bromomethane	0.185	0.202	-9.2	118	0.00
6 T	Chloroethane	0.277	0.242	12.6	92	0.00
7 T	Trichlorofluoromethane	0.720	0.717	0.4	99	0.00
8 T	Diethyl Ether	0.259	0.255	1.5	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.468	0.474	-1.3	102	0.00
10 T	Methyl Iodide	0.544	0.538	1.1	96	0.00
11 T	Tert butyl alcohol	0.124	0.117	5.6	92	0.00
12 CM	1,1-Dichloroethene	0.430	0.418	2.8#	103	0.00
13 T	Acrolein	0.093	0.095	-2.2	100	0.00
14 T	Allyl chloride	0.789	0.788	0.1	102	0.00
15 T	Acrylonitrile	0.291	0.294	-1.0	97	0.00
16 T	Acetone	0.286	0.264	7.7	93	0.00
17 T	Carbon Disulfide	0.879	0.830	5.6	102	0.00
18 T	Methyl Acetate	0.704	0.625	11.2	90	0.00
19 T	Methyl tert-butyl Ether	1.637	1.697	-3.7	102	0.00
20 T	Methylene Chloride	0.514	0.499	2.9	101	0.00
21 T	trans-1,2-Dichloroethene	0.466	0.441	5.4	100	0.00
22 T	Diisopropyl ether	1.581	1.527	3.4	106	0.00
23 T	Vinyl Acetate	1.366	1.337	2.1	106	0.00
24 P	1,1-Dichloroethane	0.889	0.842	5.3	103	0.00
25 T	2-Butanone	0.415	0.372	10.4	98	0.00
26 T	2,2-Dichloropropane	0.710	0.670	5.6	101	0.00
27 T	cis-1,2-Dichloroethene	0.562	0.551	2.0	99	0.00
28 T	Bromochloromethane	0.382	0.348	8.9	100	0.00
29 T	Tetrahydrofuran	0.267	0.236	11.6	96	0.00
30 C	Chloroform	0.988	0.900	8.9#	97	0.00
31 T	Cyclohexane	0.713	0.633	11.2	93	0.00
32 T	1,1,1-Trichloroethane	0.855	0.789	7.7	97	0.00
33 S	1,2-Dichloroethane-d4	0.645	0.570	11.6	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	107	0.00
35 S	Dibromofluoromethane	0.323	0.310	4.0	105	0.00
36 T	1,1-Dichloropropene	0.406	0.364	10.3	94	0.00
37 T	Ethyl Acetate	0.504	0.439	12.9	100	0.00
38 T	Carbon Tetrachloride	0.449	0.422	6.0	96	0.00
39 T	Methylcyclohexane	0.478	0.436	8.8	96	0.00
40 TM	Benzene	1.232	1.130	8.3	93	0.00
41 T	Methacrylonitrile	0.268	0.233	13.1	94	0.00
42 TM	1,2-Dichloroethane	0.492	0.414	15.9	80	0.00
43 T	Isopropyl Acetate	0.807	0.684	15.2	78	0.00
44 TM	Trichloroethene	0.335	0.332	0.9	105	0.00
45 C	1,2-Dichloropropane	0.331	0.296	10.6#	90	0.00
46 T	Dibromomethane	0.249	0.223	10.4	91	0.00
47 T	Bromodichloromethane	0.496	0.424	14.5	86	0.00
48 T	Methyl methacrylate	0.423	0.335	20.8	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032922\
 Data File : VX027783.D
 Acq On : 29 Mar 2022 18:44
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX032922

Quant Time: Mar 30 04:41:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	90	0.00
50 S	Toluene-d8	1.136	1.120	1.4	106	0.00
51 T	4-Methyl-2-Pentanone	0.560	0.454	18.9	79	0.00
52 CM	Toluene	0.776	0.757	2.4#	99	0.00
53 T	t-1,3-Dichloropropene	0.497	0.450	9.5	89	0.00
54 T	cis-1,3-Dichloropropene	0.523	0.480	8.2	91	0.00
55 T	1,1,2-Trichloroethane	0.344	0.328	4.7	98	0.00
56 T	Ethyl methacrylate	0.545	0.506	7.2	91	0.00
57 T	1,3-Dichloropropane	0.579	0.528	8.8	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.272	0.242	11.0	86	0.00
59 T	2-Hexanone	0.431	0.345	20.0	78	0.00
60 T	Dibromochloromethane	0.357	0.353	1.1	99	0.00
61 T	1,2-Dibromoethane	0.360	0.346	3.9	97	0.00
62 S	4-Bromofluorobenzene	0.483	0.452	6.4	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
64 T	Tetrachloroethene	0.339	0.335	1.2	105	0.00
65 PM	Chlorobenzene	0.985	0.951	3.5	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.369	0.356	3.5	100	0.00
67 C	Ethyl Benzene	1.787	1.625	9.1#	95	0.00
68 T	m/p-Xylenes	0.657	0.635	3.3	101	0.00
69 T	o-Xylene	0.642	0.636	0.9	102	0.00
70 T	Styrene	1.091	1.055	3.3	100	0.00
71 P	Bromoform	0.264	0.275	-4.2	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
73 T	Isopropylbenzene	3.723	3.398	8.7	99	0.00
74 T	N-amyl acetate	1.677	1.312	21.8	80	0.00
75 P	1,1,2,2-Tetrachloroethane	1.264	1.100	13.0	92	0.00
76 T	1,2,3-Trichloropropane	1.173	1.018	13.2	89	0.00
77 T	Bromobenzene	0.849	0.846	0.4	105	0.00
78 T	n-propylbenzene	4.448	3.861	13.2	94	0.00
79 T	2-Chlorotoluene	2.795	2.345	16.1	93	0.00
80 T	1,3,5-Trimethylbenzene	3.230	2.915	9.8	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.359	0.319	11.1	90	0.00
82 T	4-Chlorotoluene	3.207	2.700	15.8	91	0.00
83 T	tert-Butylbenzene	3.124	2.914	6.7	102	0.00
84 T	1,2,4-Trimethylbenzene	3.221	2.915	9.5	97	0.00
85 T	sec-Butylbenzene	3.855	3.532	8.4	97	0.00
86 T	p-Isopropyltoluene	3.206	3.042	5.1	100	0.00
87 T	1,3-Dichlorobenzene	1.671	1.565	6.3	104	0.00
88 T	1,4-Dichlorobenzene	1.721	1.582	8.1	103	0.00
89 T	n-Butylbenzene	2.883	2.555	11.4	93	0.00
90 T	Hexachloroethane	0.515	0.469	8.9	98	0.00
91 T	1,2-Dichlorobenzene	1.650	1.564	5.2	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.304	0.252	17.1	85	0.00
93 T	1,2,4-Trichlorobenzene	1.009	1.012	-0.3	112	0.00
94 T	Hexachlorobutadiene	0.471	0.454	3.6	111	0.00
95 T	Naphthalene	3.524	3.541	-0.5	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032922\
Data File : VX027783.D
Acq On : 29 Mar 2022 18:44
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX032922

Quant Time: Mar 30 04:41:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.989	0.994	-0.5	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6