

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032922\
 Data File : VX027779.D
 Acq On : 29 Mar 2022 17:11
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/30/2022
 Supervised By : Mahesh Dadoda 03/30/2022

Quant Time: Mar 29 18:02:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 29 17:59:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	74975	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	117793	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	107598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51683	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	49278	40.715	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	81.440%		
35) Dibromofluoromethane	5.391	113	37475	47.482	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.960%		
50) Toluene-d8	8.653	98	134059	44.725	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	89.460%		
62) 4-Bromofluorobenzene	11.085	95	57906	47.071	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	94.140%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	38990	46.362	ug/l	100
3) Chloromethane	1.294	50	31005	36.046	ug/l	99
4) Vinyl Chloride	1.374	62	32736	39.456	ug/l	99
5) Bromomethane	1.599	94	13323	39.231	ug/l	99
6) Chloroethane	1.678	64	20392	37.898	ug/l	98
7) Trichlorofluoromethane	1.886	101	56408	44.716	ug/l	97
8) Diethyl Ether	2.136	74	20236	44.552	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.331	101	36135	42.855	ug/l	95
10) Methyl Iodide	2.453	142	43282	51.959	ug/l #	89
11) Tert butyl alcohol	2.965	59	49545	223.795	ug/l	100
12) 1,1-Dichloroethene	2.319	96	31486	40.623	ug/l	95
13) Acrolein	2.233	56	36702	229.422	ug/l	95
14) Allyl chloride	2.666	41	59743	41.241	ug/l #	92
15) Acrylonitrile	3.062	53	117358	216.092	ug/l	99
16) Acetone	2.380	43	109982	212.934	ug/l	89
17) Carbon Disulfide	2.514	76	62850	31.239	ug/l	99
18) Methyl Acetate	2.703	43	53733	41.942	ug/l	91
19) Methyl tert-butyl Ether	3.111	73	129002	45.094	ug/l	97
20) Methylene Chloride	2.788	84	38267	42.306	ug/l	95
21) trans-1,2-Dichloroethene	3.093	96	34349	39.734	ug/l	91
22) Diisopropyl ether	3.763	45	111308	38.518	ug/l #	90
23) Vinyl Acetate	3.727	43	489426	194.712	ug/l	96
24) 1,1-Dichloroethane	3.611	63	63213	38.764	ug/l	100
25) 2-Butanone	4.556	43	146991	183.048	ug/l	94
26) 2,2-Dichloropropane	4.483	77	51526	40.527	ug/l	100
27) cis-1,2-Dichloroethene	4.495	96	43070	44.376	ug/l	96
28) Bromochloromethane	4.897	49	27046	41.098	ug/l	95
29) Tetrahydrofuran	5.007	42	95153	197.203	ug/l	90
30) Chloroform	5.099	83	72221	43.005	ug/l	95
31) Cyclohexane	5.477	56	52635	36.928	ug/l	95
32) 1,1,1-Trichloroethane	5.385	97	63163	42.294	ug/l	99
36) 1,1-Dichloropropene	5.690	75	48994	40.804	ug/l	98
37) Ethyl Acetate	4.715	43	55394	40.253	ug/l	99
38) Carbon Tetrachloride	5.684	117	55873	46.713	ug/l	99
39) Methylcyclohexane	7.385	83	57700	42.184	ug/l	90
40) Benzene	6.037	78	153706	46.132	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032922\
 Data File : VX027779.D
 Acq On : 29 Mar 2022 17:11
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/30/2022
 Supervised By : Mahesh Dadoda 03/30/2022

Quant Time: Mar 29 18:02:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 29 17:59:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	31412	43.448	ug/l	94
42) 1,2-Dichloroethane	6.092	62	65164	46.212	ug/l	93
43) Isopropyl Acetate	6.342	43	110297	51.645	ug/l #	90
44) Trichloroethene	7.129	130	40081	45.836	ug/l	88
45) 1,2-Dichloropropane	7.434	63	41694	47.391	ug/l	98
46) Dibromomethane	7.586	93	31093	49.201	ug/l #	88
47) Bromodichloromethane	7.824	83	62185	50.181	ug/l	94
48) Methyl methacrylate	7.696	41	55261	52.487	ug/l #	79
49) 1,4-Dioxane	7.659	88	21632	953.466	ug/l #	87
51) 4-Methyl-2-Pentanone	8.574	43	362879	247.762	ug/l	88
52) Toluene	8.720	92	96451	46.121	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	64008	52.132	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	67124	49.412	ug/l #	77
55) 1,1,2-Trichloroethane	9.153	97	42263	49.181	ug/l	96
56) Ethyl methacrylate	9.116	69	70671	50.696	ug/l #	72
57) 1,3-Dichloropropane	9.311	76	72595	47.241	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	177743	249.967	ug/l	95
59) 2-Hexanone	9.433	43	280900	244.961	ug/l	83
60) Dibromochloromethane	9.525	129	45257	53.874	ug/l	96
61) 1,2-Dibromoethane	9.610	107	44987	49.003	ug/l	98
64) Tetrachloroethene	9.275	164	37269	45.971	ug/l	94
65) Chlorobenzene	10.079	112	107403	47.337	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	41787	52.694	ug/l	99
67) Ethyl Benzene	10.195	91	201416	47.894	ug/l	99
68) m/p-Xylenes	10.305	106	147809	96.843	ug/l	90
69) o-Xylene	10.640	106	73330	47.693	ug/l	88
70) Styrene	10.659	104	124204	50.708	ug/l #	92
71) Bromoform	10.805	173	31267	54.309	ug/l #	99
73) Isopropylbenzene	10.963	105	197923	47.155	ug/l	98
74) N-amyl acetate	10.841	43	94800	47.182	ug/l #	85
75) 1,1,2,2-Tetrachloroethane	11.213	83	68404	46.100	ug/l	100
76) 1,2,3-Trichloropropane	11.244	75	65619m	46.238	ug/l	
77) Bromobenzene	11.201	156	46096	48.819	ug/l	84
78) n-propylbenzene	11.305	91	234973	47.227	ug/l	98
79) 2-Chlorotoluene	11.366	91	145319	46.179	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	172318	47.634	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.018	75	20422	48.917	ug/l #	79
82) 4-Chlorotoluene	11.457	91	170194	46.957	ug/l	95
83) tert-Butylbenzene	11.719	119	164504	49.203	ug/l	86
84) 1,2,4-Trimethylbenzene	11.756	105	172567	47.044	ug/l	95
85) sec-Butylbenzene	11.890	105	208135	47.746	ug/l	98
86) p-Isopropyltoluene	12.012	119	175016	50.765	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	86680	49.260	ug/l	96
88) 1,4-Dichlorobenzene	12.042	146	87912	48.179	ug/l	96
89) n-Butylbenzene	12.335	91	157867	48.494	ug/l	96
90) Hexachloroethane	12.542	117	27435	49.829	ug/l	90
91) 1,2-Dichlorobenzene	12.335	146	85411	50.080	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.945	75	17017	46.564	ug/l	76
93) 1,2,4-Trichlorobenzene	13.591	180	52066	52.338	ug/l	97
94) Hexachlorobutadiene	13.725	225	23486	48.638	ug/l	95
95) Naphthalene	13.780	128	192013	52.318	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	53305	53.154	ug/l	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032922\
Data File : VX027779.D
Acq On : 29 Mar 2022 17:11
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Manual Integrations
APPROVED

Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022

Quant Time: Mar 29 18:02:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 29 17:59:33 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:168 Resp: 74975
Ion Ratio Lower Upper
168 100
99 60.6 41.7 62.5




Manual Integrations
APPROVED

Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#2
Dichlorodifluoromethane
Concen: 46.362 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Tgt Ion: 85 Resp: 38990
Ion Ratio Lower Upper
85 100
87 32.2 16.2 48.6



Abundance Scan 34 (1.294 min): VX027779.D\data.ms (-29)







#3
Chloromethane
Concen: 36.046 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 50 Resp: 31000
Ion Ratio Lower Upper
50 100
52 32.6 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#4
Vinyl Chloride
Concen: 39.456 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11


Tgt Ion: 62 Resp: 32736
Ion Ratio Lower Upper
62 100
64 32.3 25.5 38.3





#5
Bromomethane
Concen: 39.231 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 94 Resp: 1332
Ion Ratio Lower Upper
94 100
96 95.2 75.3 112.9




Manual Integrations
APPROVED

Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#6
Chloroethane
Concen: 37.898 ug/l
RT: 1.678 min Scan# 97
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Tgt Ion: 64 Resp: 20392
Ion Ratio Lower Upper
64 100
66 30.7 25.3 37.9



Abundance Scan 131 (1.886 min): VX027779.D\data.ms (-12



#7

Trichlorofluoromethane

Concen: 44.716 ug/l

RT: 1.886 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:101 Resp: 56408

Ion Ratio Lower Upper

101 100





103 65.3 50.5 75.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/30/2022

Supervised By :Mahesh Dadoda 03/30/2022



#8

Diethyl Ether

Concen: 44.552 ug/l

RT: 2.136 min Scan# 172

Delta R.T. 0.000 min

Lab File: VX027779.D




Acq: 29 Mar 2022 17:11


Tgt Ion: 74 Resp: 20236

Ion Ratio Lower Upper

74 100

45 95.5 42.0 126.0









#9
1,1,2-Trichlorotrifluoroethane
Concen: 42.855 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:101 Resp: 3613
Ion Ratio Lower Upper
101 100
85 47.6 33.5 50.3
151 75.1 61.8 92.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#10
Methyl Iodide
Concen: 51.959 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11


Tgt Ion:142 Resp: 43282
Ion Ratio Lower Upper
142 100
127 51.9 33.9 50.9#
141 14.7 11.4 17.2





#11
Tert butyl alcohol
Concen: 223.795 ug/l
RT: 2.965 min Scan# 308
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 59 Resp: 4954
Ion Ratio Lower Upper
59 100
57 10.3 8.3 12.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#12
1,1-Dichloroethene
Concen: 40.623 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Tgt Ion: 96 Resp: 31486
Ion Ratio Lower Upper
96 100
61 164.5 124.6 187.0
98 61.9 51.0 76.4




Abundance Scan 188 (2.233 min): VX027779.D\data.ms (-18



m/z--> Scan 188 (2.233 min): VX027779.D\data.ms



m/z--> Scan 188 (2.233 min): VX027779.D\data.ms (-13



m/z-->

#13

Acrolein

Concen: 229.422 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Tgt Ion: 56 Resp: 36702

Ion Ratio Lower Upper

56 100

55 72.7 55.0 82.6

Instrument :

MSVOA_X

ClientSampleId :


VSTDICCC050

Manual Integrations


APPROVED

Reviewed By :John Carlone 03/30/2022


Supervised By :Mahesh Dadoda 03/30/2022




Abundance Scan 259 (2.666 min): VX027779.D\data.ms (-24



m/z--> Scan 259 (2.666 min): VX027779.D\data.ms



m/z--> Scan 259 (2.666 min): VX027779.D\data.ms (-21



m/z-->

#14

Allyl chloride

Concen: 41.241 ug/l

RT: 2.666 min Scan# 259

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11


Tgt Ion: 41 Resp: 59743


Ion Ratio Lower Upper

41 100

39 66.2 50.4 75.6


76 32.6 32.7 49.1#





#15
Acrylonitrile
Concen: 216.092 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 53 Resp: 117358
Ion Ratio Lower Upper
53 100
52 82.8 65.7 98.5
51 36.7 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#16
Acetone
Concen: 212.934 ug/l
RT: 2.380 min Scan# 212
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Tgt Ion: 43 Resp: 109982
Ion Ratio Lower Upper
43 100
58 29.5 28.8 43.2





#17
Carbon Disulfide
Concen: 31.239 ug/l
RT: 2.514 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 76 Resp: 62850
Ion Ratio Lower Upper
76 100
78 9.4 7.2 10.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#18
Methyl Acetate
Concen: 41.942 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Tgt Ion: 43 Resp: 53733
Ion Ratio Lower Upper
43 100
74 23.8 22.8 34.2





#19
Methyl tert-butyl Ether
Concen: 45.094 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 73 Resp: 12900
Ion Ratio Lower Upper
73 100
57 19.8 16.8 25.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#20
Methylene Chloride
Concen: 42.306 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Tgt Ion: 84 Resp: 38267
Ion Ratio Lower Upper
84 100
49 114.1 85.5 128.3
51 36.1 26.6 39.8
86 63.8 51.5 77.3





#21
trans-1,2-Dichloroethene
Concen: 39.734 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 96 Resp: 34349
Ion Ratio Lower Upper
96 100
61 148.0 105.4 158.2
98 62.2 50.0 75.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#22
Diisopropyl ether
Concen: 38.518 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Tgt Ion: 45 Resp: 111308
Ion Ratio Lower Upper
45 100
43 76.2 50.8 76.2#
87 33.1 26.6 39.8
59 11.8 10.5 15.7





#23
 Vinyl Acetate
 Concen: 194.712 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: VX027779.D
 Acq: 29 Mar 2022 17:11




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050




Tgt Ion: 43 Resp: 489420
 Ion Ratio Lower Upper
 43 100
 86 11.9 10.8 16.2

Manual Integrations
 APPROVED



Reviewed By :John Carlone 03/30/2022
 Supervised By :Mahesh Dadoda 03/30/2022




#24
 1,1-Dichloroethane
 Concen: 38.764 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX027779.D
 Acq: 29 Mar 2022 17:11



Tgt Ion: 63 Resp: 63213
 Ion Ratio Lower Upper
 63 100
 98 7.9 3.9 11.7
 100 4.9 2.4 7.2





#25

2-Butanone

Concen: 183.048 ug/l

RT: 4.556 min Scan# 50

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 14699

Ion Ratio Lower Upper

43 100





72 27.3 24.4 36.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/30/2022

Supervised By :Mahesh Dadoda 03/30/2022



#26

2,2-Dichloropropane

Concen: 40.527 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX027779.D




Acq: 29 Mar 2022 17:11


Tgt Ion: 77 Resp: 51526

Ion Ratio Lower Upper

77 100


97 23.8 11.8 35.4





#27
 cis-1,2-Dichloroethene
 Concen: 44.376 ug/l
 RT: 4.495 min Scan# 559
 Delta R.T. 0.000 min
 Lab File: VX027779.D
 Acq: 29 Mar 2022 17:11




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050




Tgt Ion: 96 Resp: 43070
 Ion Ratio Lower Upper
 96 100
 61 136.1 0.0 283.8
 98 63.8 0.0 129.8

Manual Integrations
 APPROVED



Reviewed By :John Carlone 03/30/2022
 Supervised By :Mahesh Dadoda 03/30/2022




#28
 Bromochloromethane
 Concen: 41.098 ug/l
 RT: 4.897 min Scan# 625
 Delta R.T. 0.000 min
 Lab File: VX027779.D
 Acq: 29 Mar 2022 17:11



Tgt Ion: 49 Resp: 27046
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 4.4
 130 93.4 70.7 106.1





#29

Tetrahydrofuran

Concen: 197.203 ug/l

RT: 5.007 min Scan# 643

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 42 Resp: 9515





Ion	Ratio	Lower	Upper
42	100		
72	47.5	44.1	66.1
71	44.7	40.5	60.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/30/2022

Supervised By :Mahesh Dadoda 03/30/2022



#30

Chloroform

Concen: 43.005 ug/l

RT: 5.099 min Scan# 658




Delta R.T. 0.000 min


Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Tgt Ion: 83 Resp: 72221


Ion	Ratio	Lower	Upper
83	100		
85	61.7	52.7	79.1





#31
Cyclohexane
Concen: 36.928 ug/l
RT: 5.477 min Scan# 720
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 56 Resp: 5263
Ion Ratio Lower Upper
56 100
69 31.8 26.7 40.1
84 90.8 77.2 115.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#32
1,1,1-Trichloroethane
Concen: 42.294 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11



Tgt Ion: 97 Resp: 63163
Ion Ratio Lower Upper
97 100
99 64.3 51.8 77.6
61 41.5 34.0 51.0









#33
1,2-Dichloroethane-d4
Concen: 40.715 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 65 Resp: 49278
Ion Ratio Lower Upper
65 100
67 50.8 0.0 107.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11


Tgt Ion:114 Resp: 117793
Ion Ratio Lower Upper
114 100
63 23.4 0.0 36.8
88 18.2 0.0 32.4





#35
Dibromofluoromethane
Concen: 47.482 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion:113 Resp: 37475
Ion Ratio Lower Upper
113 100
111 102.7 82.8 124.2
192 19.1 16.9 25.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#36
1,1-Dichloropropene
Concen: 40.804 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Tgt Ion: 75 Resp: 48994
Ion Ratio Lower Upper
75 100
110 37.7 18.3 54.8
77 32.0 24.8 37.2





#37
Ethyl Acetate
Concen: 40.253 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 43 Resp: 55394
Ion Ratio Lower Upper
43 100
61 15.1 12.6 19.0
70 12.7 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#38
Carbon Tetrachloride
Concen: 46.713 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Tgt Ion:117 Resp: 55873
Ion Ratio Lower Upper
117 100
119 95.4 77.3 115.9
121 29.5 24.2 36.2





#39

Methylcyclohexane

Concen: 42.184 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 57700





Ion	Ratio	Lower	Upper
83	100		
55	80.4	54.6	81.8
98	45.5	37.4	56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/30/2022

Supervised By :Mahesh Dadoda 03/30/2022



#40

Benzene

Concen: 46.132 ug/l

RT: 6.037 min Scan# 812




Delta R.T. 0.000 min


Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Tgt Ion: 78 Resp: 153706


Ion	Ratio	Lower	Upper
78	100		
77	23.0	19.3	28.9





#41
Methacrylonitrile
Concen: 43.448 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 41 Resp: 3141
Ion Ratio Lower Upper
41 100
39 54.0 43.0 64.4
67 77.1 70.2 105.4
52 31.5 26.0 39.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#42
1,2-Dichloroethane
Concen: 46.212 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Tgt Ion: 62 Resp: 65164
Ion Ratio Lower Upper
62 100
98 7.5 0.0 20.4





#43
Isopropyl Acetate
Concen: 51.645 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 43 Resp: 110297
Ion Ratio Lower Upper
43 100
61 18.3 18.5 27.7
87 12.4 13.0 19.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#44
Trichloroethene
Concen: 45.836 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11



Tgt Ion:130 Resp: 40081
Ion Ratio Lower Upper
130 100
95 106.0 0.0 188.8





#45

1,2-Dichloropropane

Concen: 47.391 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 41694

Ion Ratio Lower Upper

63 100





65 30.8 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/30/2022

Supervised By :Mahesh Dadoda 03/30/2022



#46

Dibromomethane

Concen: 49.201 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11




Tgt Ion: 93 Resp: 31093


Ion Ratio Lower Upper

93 100

95 82.3 65.8 98.6


174 85.0 85.4 128.2#





#47
Bromodichloromethane
Concen: 50.181 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 83 Resp: 62185
Ion Ratio Lower Upper
83 100
85 60.2 52.2 78.2
127 7.3 6.8 10.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#48
Methyl methacrylate
Concen: 52.487 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11



Tgt Ion: 41 Resp: 55261
Ion Ratio Lower Upper
41 100
69 75.6 82.5 123.7#
39 62.5 43.8 65.8





#49

1,4-Dioxane

Concen: 953.466 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 88 Resp: 2163





Ion	Ratio	Lower	Upper
88	100		
43	38.0	21.9	32.9
58	75.1	53.7	80.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/30/2022

Supervised By :Mahesh Dadoda 03/30/2022



#50

Toluene-d8

Concen: 44.725 ug/l

RT: 8.653 min Scan# 1241




Delta R.T. 0.000 min


Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Tgt Ion: 98 Resp: 134059


Ion	Ratio	Lower	Upper
98	100		
100	63.9	52.6	78.8





#51
4-Methyl-2-Pentanone
Concen: 247.762 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 43 Resp: 362879
Ion Ratio Lower Upper
43 100
58 37.1 36.1 54.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#52
Toluene
Concen: 46.121 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11



Tgt Ion: 92 Resp: 96451
Ion Ratio Lower Upper
92 100
91 174.2 137.5 206.3





#53

t-1,3-Dichloropropene

Concen: 52.132 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 75 Resp: 6400

Ion Ratio Lower Upper

75 100





77 30.4 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/30/2022

Supervised By :Mahesh Dadoda 03/30/2022



#54

cis-1,3-Dichloropropene

Concen: 49.412 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11




Tgt Ion: 75 Resp: 67124


Ion Ratio Lower Upper

75 100

77 30.4 24.8 37.2


39 61.3 29.6 44.4#





#55
1,1,2-Trichloroethane
Concen: 49.181 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050






Tgt Ion: 97 Resp: 4226


Ion	Ratio	Lower	Upper
97	100		
83	87.0	66.2	99.4
85	55.1	42.9	64.3
99	64.2	49.0	73.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022






#56
Ethyl methacrylate
Concen: 50.696 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11



Tgt Ion: 69 Resp: 70671

Ion	Ratio	Lower	Upper
69	100		
41	73.3	43.8	65.6#
39	45.3	22.2	33.2#









#57
1,3-Dichloropropane
Concen: 47.241 ug/l
RT: 9.311 min Scan# 1134
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 76 Resp: 7259
Ion Ratio Lower Upper
76 100
78 31.9 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#58
2-Chloroethyl Vinyl ether
Concen: 249.967 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Tgt Ion: 63 Resp: 177743
Ion Ratio Lower Upper
63 100
106 27.2 23.9 35.9





#59

2-Hexanone

Concen: 244.961 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 280900

Ion Ratio Lower Upper

43 100





58 50.7 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/30/2022

Supervised By :Mahesh Dadoda 03/30/2022



#60

Dibromochloromethane

Concen: 53.874 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. 0.000 min

Lab File: VX027779.D




Acq: 29 Mar 2022 17:11


Tgt Ion:129 Resp: 45257

Ion Ratio Lower Upper

129 100


127 77.4 40.3 120.8





#61
1,2-Dibromoethane
Concen: 49.003 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion:107 Resp: 4498
Ion Ratio Lower Upper
107 100
109 95.3 74.4 111.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#62
4-Bromofluorobenzene
Concen: 47.071 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Tgt Ion: 95 Resp: 57906
Ion Ratio Lower Upper
95 100
174 67.8 0.0 163.0
176 66.1 0.0 155.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:117 Resp: 107598
Ion Ratio Lower Upper
117 100
82 62.2 41.8 62.8
119 32.9 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#64
Tetrachloroethene
Concen: 45.971 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Tgt Ion:164 Resp: 37269
Ion Ratio Lower Upper
164 100
166 121.9 101.2 151.8
129 103.7 73.0 109.4
131 95.2 74.2 111.2









#65
Chlorobenzene
Concen: 47.337 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:112 Resp: 10740
Ion Ratio Lower Upper
112 100
114 31.9 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 52.694 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11


Tgt Ion:131 Resp: 41787
Ion Ratio Lower Upper
131 100
133 95.3 48.3 144.8
119 64.2 32.6 98.0





#67
Ethyl Benzene
Concen: 47.894 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 91 Resp: 201410
Ion Ratio Lower Upper
91 100
106 30.2 24.6 36.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#68
m/p-Xylenes
Concen: 96.843 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11



Tgt Ion:106 Resp: 147809
Ion Ratio Lower Upper
106 100
91 215.5 160.4 240.6









#69
o-Xylene
Concen: 47.693 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:106 Resp: 73330
Ion Ratio Lower Upper
106 100
91 231.3 106.1 318.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#70
Styrene
Concen: 50.708 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11


Tgt Ion:104 Resp: 124204
Ion Ratio Lower Upper
104 100
78 59.3 38.6 57.8#
103 55.8 43.8 65.8





#71
Bromoform
Concen: 54.309 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:173 Resp: 3126
Ion Ratio Lower Upper
173 100
175 48.9 24.1 72.2
254 0.0 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Tgt Ion:152 Resp: 51683
Ion Ratio Lower Upper
152 100
115 91.5 37.5 112.7
150 175.1 0.0 338.2









#73
Isopropylbenzene
Concen: 47.155 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 197923
Ion Ratio Lower Upper
105 100
120 26.1 13.5 40.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#74
N-ethyl acetate
Concen: 47.182 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11


Tgt Ion: 43 Resp: 94800
Ion Ratio Lower Upper
43 100
70 39.0 42.5 63.7#
55 23.5 23.1 34.7
61 22.1 22.8 34.2#





#75
1,1,2,2-Tetrachloroethane
Concen: 46.100 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 83 Resp: 68404
Ion Ratio Lower Upper
83 100
131 10.6 5.1 15.3
85 65.4 32.6 97.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#76
1,2,3-Trichloropropane
Concen: 46.238 ug/l m
RT: 11.244 min Scan# 1666
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Tgt Ion: 75 Resp: 65619
Ion Ratio Lower Upper
75 100
77 37.2 19.4 58.2





#77
Bromobenzene
Concen: 48.819 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion:156 Resp: 46090
Ion Ratio Lower Upper
156 100
77 187.7 78.5 235.3
158 95.8 50.2 150.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#78
n-propylbenzene
Concen: 47.227 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11



Tgt Ion: 91 Resp: 234973
Ion Ratio Lower Upper
91 100
120 22.2 11.6 34.8









#79
2-Chlorotoluene
Concen: 46.179 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 145319
Ion Ratio Lower Upper
91 100
126 31.7 17.3 51.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#80
1,3,5-Trimethylbenzene
Concen: 47.634 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11


Tgt Ion:105 Resp: 172318
Ion Ratio Lower Upper
105 100
120 46.2 24.3 72.9





#81
trans-1,4-Dichloro-2-butene
Concen: 48.917 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 75 Resp: 2042
Ion Ratio Lower Upper
75 100
53 114.9 69.0 103.4
89 47.3 37.0 55.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#82
4-Chlorotoluene
Concen: 46.957 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11



Tgt Ion: 91 Resp: 170194
Ion Ratio Lower Upper
91 100
126 27.2 14.9 44.5









#83
tert-Butylbenzene
Concen: 49.203 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:119 Resp: 164504
Ion Ratio Lower Upper
119 100
91 67.5 26.9 80.6
134 22.1 11.4 34.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#84
1,2,4-Trimethylbenzene
Concen: 47.044 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11


Tgt Ion:105 Resp: 172567
Ion Ratio Lower Upper
105 100
120 42.1 22.6 67.8





#85
 sec-Butylbenzene
 Concen: 47.746 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX027779.D
 Acq: 29 Mar 2022 17:11




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050




Tgt Ion:105 Resp: 20813
 Ion Ratio Lower Upper
 105 100
 134 19.6 10.2 30.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 03/30/2022
 Supervised By :Mahesh Dadoda 03/30/2022




#86
 p-Isopropyltoluene
 Concen: 50.765 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX027779.D
 Acq: 29 Mar 2022 17:11



Tgt Ion:119 Resp: 175016
 Ion Ratio Lower Upper
 119 100
 134 25.7 13.0 39.0
 91 28.6 11.3 33.9




Abundance Scan 1785 (11.969 min): VX027779.D\data.ms (-




m/z-->

Abundance Scan 1785 (11.969 min): VX027779.D\data.ms




m/z-->

Abundance Scan 1785 (11.969 min): VX027779.D\data.ms (-




m/z-->

Abundance Scan 1797 (12.042 min): VX027779.D\data.ms (-




m/z-->

Abundance Scan 1797 (12.042 min): VX027779.D\data.ms



m/z-->

Abundance Scan 1797 (12.042 min): VX027779.D\data.ms (-



m/z-->

#87

1,3-Dichlorobenzene

Concen: 49.260 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Tgt Ion:146 Resp: 86680

Ion Ratio Lower Upper

146 100

111 45.0 19.7 59.1

148 62.9 31.6 94.7

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050


Manual Integrations

APPROVED

Reviewed By :John Carlone 03/30/2022

Supervised By :Mahesh Dadoda 03/30/2022

Abundance



Time-->

#88

1,4-Dichlorobenzene

Concen: 48.179 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Tgt Ion:146 Resp: 87912


Ion Ratio Lower Upper

146 100


111 43.2 19.2 57.6

148 62.7 32.3 96.9

Abundance



Time-->



#89

n-Butylbenzene

Concen: 48.494 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX027779.D

Acq: 29 Mar 2022 17:11

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 15786

Ion Ratio Lower Upper

91 100

92 52.3 27.8 83.3





134 24.3 13.3 39.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/30/2022

Supervised By :Mahesh Dadoda 03/30/2022



#90

Hexachloroethane

Concen: 49.829 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX027779.D




Acq: 29 Mar 2022 17:11


Tgt Ion:117 Resp: 27435

Ion Ratio Lower Upper

117 100


201 70.0 39.5 118.5





#91
1,2-Dichlorobenzene
Concen: 50.080 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion:146 Resp: 8541
Ion Ratio Lower Upper
146 100
111 45.7 21.0 63.0
148 62.7 32.6 97.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 46.564 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Tgt Ion: 75 Resp: 17017
Ion Ratio Lower Upper
75 100
155 70.6 47.0 141.2
157 89.7 57.9 173.6





#93
1,2,4-Trichlorobenzene
Concen: 52.338 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion:180 Resp: 52060
Ion Ratio Lower Upper
180 100
182 94.0 47.7 143.1
145 34.6 15.2 45.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022




#94
Hexachlorobutadiene
Concen: 48.638 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Tgt Ion:225 Resp: 23486
Ion Ratio Lower Upper
225 100
223 64.7 34.2 102.5
227 63.0 33.9 101.7





#95
Naphthalene
Concen: 52.318 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion:128 Resp: 19201
Ion Ratio Lower Upper
128 100
127 12.9 10.2 15.4
129 10.7 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 03/30/2022
Supervised By :Mahesh Dadoda 03/30/2022



#96
1,2,3-Trichlorobenzene
Concen: 53.154 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX027779.D
Acq: 29 Mar 2022 17:11



Tgt Ion:180 Resp: 53305
Ion Ratio Lower Upper
180 100
182 92.4 50.0 149.8
145 35.7 16.2 48.5

