

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032922\
 Data File : VX027793.D
 Acq On : 29 Mar 2022 23:03
 Operator : JC/MD
 Sample : N2041-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE117D2-20220317

Quant Time: Mar 30 05:16:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	81859	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	145580	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120561	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51265	48.537	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.080%
35) Dibromofluoromethane	5.391	113	48370	51.475	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.940%
50) Toluene-d8	8.653	98	162340	49.101	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.200%
62) 4-Bromofluorobenzene	11.079	95	62956	44.808	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.620%


Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032922\
Data File : VX027793.D
Acq On : 29 Mar 2022 23:03
Operator : JC/MD
Sample : N2041-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE117D2-20220317


Quant Time: Mar 30 05:16:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 713
Delta R.T. -0.000 min
Lab File: VX027793.D
Acq: 29 Mar 2022 23:03

Instrument :
MSVOA_X
ClientSampleId :
RE117D2-20220317







Tgt Ion:168 Resp: 81859
Ion Ratio Lower Upper
168 100
99 65.6 41.7 62.5#



#33
1,2-Dichloroethane-d4
Concen: 48.537 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX027793.D
Acq: 29 Mar 2022 23:03

Tgt Ion: 65 Resp: 51265
Ion Ratio Lower Upper
65 100
67 51.9 0.0 107.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. 0.000 min
Lab File: VX027793.D
Acq: 29 Mar 2022 23:03




Instrument :
MSVOA_X
ClientSampleId :
RE117D2-20220317


Tgt Ion:114 Resp: 145580
Ion Ratio Lower Upper
114 100
63 19.0 0.0 36.8
88 16.3 0.0 32.4



#35
Dibromofluoromethane
Concen: 51.475 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX027793.D
Acq: 29 Mar 2022 23:03

Tgt Ion:113 Resp: 48370
Ion Ratio Lower Upper
113 100
111 100.9 82.8 124.2
192 18.3 16.9 25.3









#50
Toluene-d8
Concen: 49.101 ug/l
RT: 8.653 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX027793.D
Acq: 29 Mar 2022 23:03




Instrument :
MSVOA_X
ClientSampleId :
RE117D2-20220317


Tgt Ion: 98 Resp: 162340
Ion Ratio Lower Upper
98 100
100 64.4 52.6 78.8



#62
4-Bromofluorobenzene
Concen: 44.808 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.006 min
Lab File: VX027793.D
Acq: 29 Mar 2022 23:03

Tgt Ion: 95 Resp: 62956
Ion Ratio Lower Upper
95 100
174 67.9 0.0 163.0
176 64.8 0.0 155.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX027793.D
Acq: 29 Mar 2022 23:03

Instrument :
MSVOA_X
ClientSampleId :
RE117D2-20220317

Tgt Ion:117 Resp: 120561
Ion Ratio Lower Upper
117 100
82 62.7 41.8 62.8
119 31.3 24.8 37.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX027793.D
Acq: 29 Mar 2022 23:03

Tgt Ion:152 Resp: 49954
Ion Ratio Lower Upper
152 100
115 67.0 37.5 112.7
150 158.5 0.0 338.2

