

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032924\
 Data File : VX040835.D
 Acq On : 29 Mar 2024 11:31
 Operator : JC/MD
 Sample : VX0329WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0329WBL01

Quant Time: Mar 29 22:58:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032824W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 29 05:24:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	3412	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	28528	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	30953	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	18600	31.807	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	106.033%
60) 4-Bromofluorobenzene	11.079	95	21560	31.886	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	106.300%
63) Toluene-d8	8.647	98	40077	33.027	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	110.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032924\
Data File : VX040835.D
Acq On : 29 Mar 2024 11:31
Operator : JC/MD
Sample : VX0329WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0329WBL01

Quant Time: Mar 29 22:58:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032824W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 29 05:24:11 2024
Response via : Initial Calibration

