

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032924\
 Data File : VX040837.D
 Acq On : 29 Mar 2024 13:56
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 29 22:59:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032824W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 29 05:24:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	3223	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	25625	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	31586	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	16946	30.678	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.267%
60) 4-Bromofluorobenzene	11.079	95	18078	26.200	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	87.333%
63) Toluene-d8	8.647	98	37143	29.996	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032924\
Data File : VX040837.D
Acq On : 29 Mar 2024 13:56
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 29 22:59:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032824W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 29 05:24:11 2024
Response via : Initial Calibration

