

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX033020\
 Data File : VX015428.D
 Acq On : 30 Mar 2020 15:19
 Operator : JC/SP
 Sample : L2021-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-B

Quant Time: Mar 31 08:51:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1153787	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.84	114	1879524	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1781058	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	933794	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	766911	47.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.26%	
35) Dibromofluoromethane	5.47	113	596632	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
50) Toluene-d8	8.70	98	2286920	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
62) 4-Bromofluorobenzene	11.12	95	967987	53.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.70%	
Target Compounds						
					Qvalue	
16) Acetone	2.42	43	21934	3.137	ug/l	98
17) Carbon Disulfide	2.55	76	17610	0.530	ug/l #	90
19) Methyl tert-butyl Ether	3.17	73	19864	0.470	ug/l #	85
86) p-Isopropyltoluene	12.05	119	19767	0.361	ug/l #	79
89) n-Butylbenzene	12.37	91	15240	0.308	ug/l	100
93) 1,2,4-Trichlorobenzene	13.63	180	8083	0.357	ug/l	93
94) Hexachlorobutadiene	13.77	225	4438	0.412	ug/l	94
95) Naphthalene	13.82	128	20938	0.304	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX033020\
Data File : VX015428.D
Acq On : 30 Mar 2020 15:19
Operator : JC/SP
Sample : L2021-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-B

Quant Time: Mar 31 08:51:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

