

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX033020\
 Data File : VX015434.D
 Acq On : 30 Mar 2020 17:35
 Operator : JC/SP
 Sample : L2022-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-1

Quant Time: Mar 31 09:25:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1104058	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1822696	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1737422	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	938137	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	765427	49.16	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	98.32%	
35) Dibromofluoromethane	5.47	113	575669	49.41	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	98.82%	
50) Toluene-d8	8.70	98	2235693	52.12	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	104.24%	
62) 4-Bromofluorobenzene	11.12	95	968921	55.58	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	111.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.39	62	74414	6.542	ug/l	98
12) 1,1-Dichloroethene	2.36	96	581258	55.764	ug/l	94
16) Acetone	2.42	43	53711	8.029	ug/l	96
24) 1,1-Dichloroethane	3.67	63	462926	20.373	ug/l	98
27) cis-1,2-Dichloroethene	4.57	96	113280	8.464	ug/l	91
39) Methylcyclohexane	7.45	83	6842	0.353	ug/l #	70
40) Benzene	6.12	78	27523	0.559	ug/l	95
42) 1,2-Dichloroethane	6.17	62	58105	2.940	ug/l	98
44) Trichloroethene	7.20	130	10940	0.815	ug/l	88
49) 1,4-Dioxane	7.71	88	248155	721.459	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX033020\
Data File : VX015434.D
Acq On : 30 Mar 2020 17:35
Operator : JC/SP
Sample : L2022-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-1

Quant Time: Mar 31 09:25:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

