

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX033020\
 Data File : VX015435.D
 Acq On : 30 Mar 2020 17:58
 Operator : JC/SP
 Sample : L2022-05 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-1

Quant Time: Mar 31 09:28:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1119261	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1819067	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1766691	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	944060	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	765612	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
35) Dibromofluoromethane	5.47	113	579764	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
50) Toluene-d8	8.70	98	2214046	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
62) 4-Bromofluorobenzene	11.12	95	961301	55.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.50%	
Target Compounds						
12) 1,1-Dichloroethene	2.36	96	28795	2.725	ug/l	94
24) 1,1-Dichloroethane	3.67	63	22557	0.979	ug/l	94
27) cis-1,2-Dichloroethene	4.57	96	6499	0.479	ug/l	85
49) 1,4-Dioxane	7.72	88	13836	40.306	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX033020\
Data File : VX015435.D
Acq On : 30 Mar 2020 17:58
Operator : JC/SP
Sample : L2022-05 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-1

Quant Time: Mar 31 09:28:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

