

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX033020\
 Data File : VX015438.D
 Acq On : 30 Mar 2020 19:06
 Operator : JC/SP
 Sample : L2048-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-3

Quant Time: Mar 31 09:40:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1140154	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1873794	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1770899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	960886	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	777315	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
35) Dibromofluoromethane	5.47	113	580357	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
50) Toluene-d8	8.70	98	2246525	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
62) 4-Bromofluorobenzene	11.12	95	962724	53.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.44%	
Target Compounds						
16) Acetone	2.42	43	18817	2.724	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX033020\
Data File : VX015438.D
Acq On : 30 Mar 2020 19:06
Operator : JC/SP
Sample : L2048-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-3

Quant Time: Mar 31 09:40:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

