

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX033020\
 Data File : VX015439.D
 Acq On : 30 Mar 2020 19:28
 Operator : JC/SP
 Sample : L2048-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-4

Quant Time: Mar 31 12:45:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1102419	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1783941	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1709856	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	900857	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	740531	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
35) Dibromofluoromethane	5.47	113	565539	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
50) Toluene-d8	8.70	98	2165481	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
62) 4-Bromofluorobenzene	11.12	95	940715	55.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.26%	
Target Compounds						
16) Acetone	2.42	43	11459	1.715	ug/l #	86
65) Chlorobenzene	10.12	112	7270	0.213	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX033020\
Data File : VX015439.D
Acq On : 30 Mar 2020 19:28
Operator : JC/SP
Sample : L2048-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-4

Quant Time: Mar 31 12:45:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

