

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021630.D
 Acq On : 30 Mar 2021 15:29
 Operator : JC/MD
 Sample : M1867-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWBR1-102-115-032921

Quant Time: Mar 31 04:58:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	59292	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	107930	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	112874	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	48895	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	51168	51.87	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.74%	
35) Dibromofluoromethane	5.470	113	39106	51.01	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.02%	
50) Toluene-d8	8.694	98	137382	50.29	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.58%	
62) 4-Bromofluorobenzene	11.118	95	56300	53.01	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.02%	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.352	96	1282	2.15	ug/l #	91
16) Acetone	2.423	43	2166	5.81	ug/l	93
21) trans-1,2-Dichloroethene	3.146	96	2156	3.28	ug/l #	92
27) cis-1,2-Dichloroethene	4.570	96	17652	23.16	ug/l	96
30) Chloroform	5.164	83	1190	0.89	ug/l #	70
40) Benzene	6.117	78	2078	0.69	ug/l	91
44) Trichloroethene	7.188	130	61232	77.27	ug/l	94
52) Toluene	8.765	92	27319	14.73	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021630.D
Acq On : 30 Mar 2021 15:29
Operator : JC/MD
Sample : M1867-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWBR1-102-115-032921

Quant Time: Mar 31 04:58:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

