

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021632.D
 Acq On : 30 Mar 2021 16:11
 Operator : JC/MD
 Sample : M1849-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 31 04:59:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	60983	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	109994	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	113306	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	48718	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	51979	51.23	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.46%	
35) Dibromofluoromethane	5.470	113	40349	51.64	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.28%	
50) Toluene-d8	8.694	98	139864	50.23	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.46%	
62) 4-Bromofluorobenzene	11.118	95	55282	51.07	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021632.D
Acq On : 30 Mar 2021 16:11
Operator : JC/MD
Sample : M1849-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 31 04:59:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

