

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021636.D
 Acq On : 30 Mar 2021 17:39
 Operator : JC/MD
 Sample : M1843-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20210325

Quant Time: Mar 31 05:00:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	58675	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	106269	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	109330	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	42128	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	49783	50.99	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.98%	
35) Dibromofluoromethane	5.464	113	39426	52.23	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.46%	
50) Toluene-d8	8.694	98	135258	50.28	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.56%	
62) 4-Bromofluorobenzene	11.118	95	51221	48.98	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.96%	
Target Compounds						
16) Acetone	2.423	43	6047	16.38	ug/l	# 88
20) Methylene Chloride	2.828	84	1015	1.41	ug/l	# 89
52) Toluene	8.765	92	19572	10.72	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021636.D
Acq On : 30 Mar 2021 17:39
Operator : JC/MD
Sample : M1843-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20210325

Quant Time: Mar 31 05:00:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

