

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021651.D
 Acq On : 31 Mar 2021 00:17
 Operator : JC/MD
 Sample : M1843-09DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP07-20210324DL

Quant Time: Mar 31 06:11:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	55737	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	101776	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	102716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	42485	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	48338	52.12	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.24%
35) Dibromofluoromethane	5.470	113	36898	51.04	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.08%
50) Toluene-d8	8.694	98	125803	48.83	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.66%
62) 4-Bromofluorobenzene	11.118	95	50328	50.25	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.50%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.370	101	851	1.48	ug/l	94
12) 1,1-Dichloroethene	2.352	96	462	0.82	ug/l	93
27) cis-1,2-Dichloroethene	4.558	96	341	0.48	ug/l	76
38) Carbon Tetrachloride	5.753	117	402	0.38	ug/l #	72
44) Trichloroethene	7.188	130	33553	44.90	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021651.D
Acq On : 31 Mar 2021 00:17
Operator : JC/MD
Sample : M1843-09DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP07-20210324DL

Quant Time: Mar 31 06:11:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

