

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021652.D
 Acq On : 31 Mar 2021 00:40
 Operator : JC/MD
 Sample : M1843-03DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D2-20210324DL

Quant Time: Mar 31 06:11:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	62529	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	113064	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	113776	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	48060	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	53537	51.46	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.92%		
35) Dibromofluoromethane	5.464	113	41322	51.45	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.90%		
50) Toluene-d8	8.694	98	142728	49.87	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.74%		
62) 4-Bromofluorobenzene	11.118	95	54803	49.25	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 98.50%		
Target Compounds						
12) 1,1-Dichloroethene	2.352	96	255	0.40	ug/l	94
27) cis-1,2-Dichloroethene	4.570	96	7387	9.19	ug/l	97
44) Trichloroethene	7.188	130	64897	78.17	ug/l	92
64) Tetrachloroethene	9.318	164	4523	5.93	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021652.D
Acq On : 31 Mar 2021 00:40
Operator : JC/MD
Sample : M1843-03DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D2-20210324DL

Quant Time: Mar 31 06:11:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

