

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021653.D
 Acq On : 31 Mar 2021 01:04
 Operator : JC/MD
 Sample : M1843-12DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D1-20210325DL

Quant Time: Mar 31 06:11:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	59170	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	105681	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	105548	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	43147	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	50934	51.74	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.48%	
35) Dibromofluoromethane	5.464	113	38883	51.80	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.60%	
50) Toluene-d8	8.694	98	133947	50.07	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.14%	
62) 4-Bromofluorobenzene	11.118	95	49972	48.05	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.10%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.187	85	285	0.44	ug/l	83
9) 1,1,2-Trichlorotrifluo...	2.364	101	2585	4.24	ug/l	96
12) 1,1-Dichloroethene	2.358	96	628	1.05	ug/l	# 88
27) cis-1,2-Dichloroethene	4.564	96	514	0.68	ug/l	90
30) Chloroform	5.164	83	866	0.65	ug/l	93
38) Carbon Tetrachloride	5.747	117	616	0.55	ug/l	89
44) Trichloroethene	7.188	130	54532	70.28	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021653.D
Acq On : 31 Mar 2021 01:04
Operator : JC/MD
Sample : M1843-12DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D1-20210325DL

Quant Time: Mar 31 06:11:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

