

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021654.D
 Acq On : 31 Mar 2021 01:27
 Operator : JC/MD
 Sample : M1843-14DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20210325DL

Quant Time: Mar 31 06:11:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	57966	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	107839	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	111834	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	52296	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	50882	52.76	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 105.52%		
35) Dibromofluoromethane	5.464	113	39726	51.86	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.72%		
50) Toluene-d8	8.694	98	137531	50.38	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 100.76%		
62) 4-Bromofluorobenzene	11.118	95	57056	53.76	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 107.52%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.187	85	181	0.28	ug/l	91
9) 1,1,2-Trichlorotrifluo...	2.364	101	3571	5.99	ug/l	97
12) 1,1-Dichloroethene	2.358	96	367	0.63	ug/l #	77
27) cis-1,2-Dichloroethene	4.564	96	813	1.09	ug/l	94
30) Chloroform	5.170	83	829	0.63	ug/l #	80
38) Carbon Tetrachloride	5.758	117	690	0.61	ug/l #	67
44) Trichloroethene	7.188	130	74411	93.98	ug/l	98
55) 1,1,2-Trichloroethane	9.194	97	239	0.31	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021654.D
Acq On : 31 Mar 2021 01:27
Operator : JC/MD
Sample : M1843-14DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D1-20210325DL

Quant Time: Mar 31 06:11:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

