

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021657.D
 Acq On : 31 Mar 2021 02:37
 Operator : JC/MD
 Sample : M1843-07DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137-20210324DL

Quant Time: Mar 31 06:12:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	59240	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	109216	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	114388	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	52640	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	52151	52.91	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.82%	
35) Dibromofluoromethane	5.470	113	39515	50.94	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.88%	
50) Toluene-d8	8.694	98	138096	49.95	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.90%	
62) 4-Bromofluorobenzene	11.118	95	56518	52.58	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.16%	
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.370	101	525	0.86	ug/l	89
44) Trichloroethene	7.188	130	49602	61.85	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021657.D
Acq On : 31 Mar 2021 02:37
Operator : JC/MD
Sample : M1843-07DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137-20210324DL

Quant Time: Mar 31 06:12:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

