

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021659.D
 Acq On : 31 Mar 2021 03:24
 Operator : JC/MD
 Sample : M1842-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW1-7-20210323

Quant Time: Mar 31 06:13:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	58878	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	107155	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	112267	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	51714	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	51076	52.14	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.28%
35) Dibromofluoromethane	5.464	113	38948	51.17	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.34%
50) Toluene-d8	8.694	98	136002	50.14	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.28%
62) 4-Bromofluorobenzene	11.118	95	55867	52.98	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.96%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021659.D
Acq On : 31 Mar 2021 03:24
Operator : JC/MD
Sample : M1842-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW1-7-20210323

Quant Time: Mar 31 06:13:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

