

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021661.D
 Acq On : 31 Mar 2021 04:11
 Operator : JC/MD
 Sample : M1842-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D1-20210323

Quant Time: Mar 31 06:13:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	58146	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	109673	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	111912	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	50853	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	51251	52.97	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.94%	
35) Dibromofluoromethane	5.458	113	39922	51.25	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.50%	
50) Toluene-d8	8.694	98	139265	50.17	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.34%	
62) 4-Bromofluorobenzene	11.118	95	55712	51.62	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.24%	
Target Compounds						
						Qvalue
44) Trichloroethene	7.188	130	39944	49.60	ug/l	96
64) Tetrachloroethene	9.318	164	2129	2.84	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021661.D
Acq On : 31 Mar 2021 04:11
Operator : JC/MD
Sample : M1842-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D1-20210323

Quant Time: Mar 31 06:13:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

