

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021667.D
 Acq On : 31 Mar 2021 06:31
 Operator : JC/MD
 Sample : M1842-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D1-20210323

Quant Time: Mar 31 06:54:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	58387	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	106024	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	110976	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	50326	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	52051	53.58	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.16%	
35) Dibromofluoromethane	5.464	113	39912	53.00	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.00%	
50) Toluene-d8	8.694	98	137927	51.39	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.78%	
62) 4-Bromofluorobenzene	11.118	95	55633	53.32	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.64%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.370	101	3912	6.51	ug/l	92
12) 1,1-Dichloroethene	2.352	96	2396	4.07	ug/l	96
16) Acetone	2.417	43	1608	4.38	ug/l #	86
27) cis-1,2-Dichloroethene	4.558	96	2054	2.74	ug/l	95
30) Chloroform	5.182	83	3307	2.50	ug/l	93
44) Trichloroethene	7.188	130	279321	358.81	ug/l	94
55) 1,1,2-Trichloroethane	9.194	97	1366	1.78	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021667.D
Acq On : 31 Mar 2021 06:31
Operator : JC/MD
Sample : M1842-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D1-20210323

Quant Time: Mar 31 06:54:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

