

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021668.D
 Acq On : 31 Mar 2021 06:55
 Operator : JC/MD
 Sample : M1842-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D3-20210323

Quant Time: Mar 31 07:24:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	60385	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	110821	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	117414	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	52706	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	52938	52.69	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.38%
35) Dibromofluoromethane	5.464	113	40617	51.60	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.20%
50) Toluene-d8	8.694	98	142530	50.81	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.62%
62) 4-Bromofluorobenzene	11.118	95	57412	52.64	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.28%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.363	101	3834	6.17	ug/l	93
16) Acetone	2.422	43	2031	5.35	ug/l	94
44) Trichloroethene	7.188	130	409008	502.65	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021668.D
Acq On : 31 Mar 2021 06:55
Operator : JC/MD
Sample : M1842-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D3-20210323

Quant Time: Mar 31 07:24:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

