

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021618.D
 Acq On : 30 Mar 2021 10:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 31 04:55:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.550	0.507	7.8	91	0.00
3 P	Chloromethane	0.702	0.580	17.4	95	0.00
4 C	Vinyl Chloride	0.562	0.543	3.4#	95	0.00
5 T	Bromomethane	0.226	0.241	-6.6	109	0.00
6 T	Chloroethane	0.357	0.336	5.9	94	0.00
7 T	Trichlorofluoromethane	0.914	0.900	1.5	95	0.00
8 T	Diethyl Ether	0.317	0.302	4.7	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.515	0.505	1.9	99	0.00
10 T	Methyl Iodide	0.362	0.421	-16.3	116	0.00
11 T	Tert butyl alcohol	0.138	0.117	15.2	87	0.00
12 CM	1,1-Dichloroethene	0.504	0.468	7.1#	97	0.00
13 T	Acrolein	0.077	0.067	13.0	100	0.00
14 T	Allyl chloride	1.023	0.941	8.0	92	0.00
15 T	Acrylonitrile	0.327	0.300	8.3	90	0.00
16 T	Acetone	0.315	0.325	-3.2	106	0.00
17 T	Carbon Disulfide	1.415	1.160	18.0	85	0.00
18 T	Methyl Acetate	0.909	0.908	0.1	96	0.00
19 T	Methyl tert-butyl Ether	1.882	1.812	3.7	95	0.00
20 T	Methylene Chloride	0.614	0.564	8.1	94	0.00
21 T	trans-1,2-Dichloroethene	0.553	0.508	8.1	92	0.00
22 T	Diisopropyl ether	1.987	1.912	3.8	94	0.00
23 T	Vinyl Acetate	1.681	1.590	5.4	91	0.00
24 P	1,1-Dichloroethane	1.071	1.027	4.1	95	0.00
25 T	2-Butanone	0.477	0.448	6.1	93	0.00
26 T	2,2-Dichloropropane	0.928	0.906	2.4	96	0.00
27 T	cis-1,2-Dichloroethene	0.643	0.610	5.1	94	0.00
28 T	Bromochloromethane	0.536	0.479	10.6	94	0.00
29 T	Tetrahydrofuran	0.309	0.276	10.7	89	0.00
30 C	Chloroform	1.131	1.104	2.4#	96	0.00
31 T	Cyclohexane	0.915	0.861	5.9	91	0.00
32 T	1,1,1-Trichloroethane	1.002	0.983	1.9	96	0.00
33 S	1,2-Dichloroethane-d4	0.832	0.747	10.2	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.355	0.345	2.8	97	0.00
36 T	1,1-Dichloropropene	0.490	0.497	-1.4	96	0.00
37 T	Ethyl Acetate	0.566	0.523	7.6	90	0.00
38 T	Carbon Tetrachloride	0.526	0.559	-6.3	97	0.00
39 T	Methylcyclohexane	0.543	0.556	-2.4	93	0.00
40 TM	Benzene	1.391	1.419	-2.0	94	0.00
41 T	Methacrylonitrile	0.329	0.314	4.6	91	0.00
42 TM	1,2-Dichloroethane	0.603	0.619	-2.7	95	0.00
43 T	Isopropyl Acetate	0.920	0.894	2.8	90	0.00
44 TM	Trichloroethene	0.367	0.377	-2.7	97	0.00
45 C	1,2-Dichloropropane	0.377	0.391	-3.7#	95	0.00
46 T	Dibromomethane	0.275	0.282	-2.5	94	0.00
47 T	Bromodichloromethane	0.559	0.581	-3.9	95	0.00
48 T	Methyl methacrylate	0.491	0.489	0.4	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
 Data File : VX021618.D
 Acq On : 30 Mar 2021 10:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 31 04:55:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	89	0.00
50 S	Toluene-d8	1.266	1.233	2.6	94	0.00
51 T	4-Methyl-2-Pentanone	0.566	0.568	-0.4	90	0.00
52 CM	Toluene	0.859	0.908	-5.7#	95	0.00
53 T	t-1,3-Dichloropropene	0.564	0.596	-5.7	95	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.636	-5.8	95	0.00
55 T	1,1,2-Trichloroethane	0.363	0.387	-6.6	97	0.00
56 T	Ethyl methacrylate	0.548	0.566	-3.3	90	0.00
57 T	1,3-Dichloropropane	0.621	0.647	-4.2	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.307	0.323	-5.2	94	0.00
59 T	2-Hexanone	0.425	0.435	-2.4	92	0.00
60 T	Dibromochloromethane	0.406	0.441	-8.6	97	0.00
61 T	1,2-Dibromoethane	0.390	0.401	-2.8	92	0.00
62 S	4-Bromofluorobenzene	0.492	0.485	1.4	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.335	0.380	-13.4	104	0.00
65 PM	Chlorobenzene	1.004	1.036	-3.2	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.406	-5.5	96	0.00
67 C	Ethyl Benzene	1.831	1.942	-6.1#	96	0.00
68 T	m/p-Xylenes	0.652	0.700	-7.4	96	0.00
69 T	o-Xylene	0.632	0.669	-5.9	96	0.00
70 T	Styrene	1.086	1.156	-6.4	96	0.00
71 P	Bromoform	0.313	0.323	-3.2	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.513	3.661	-4.2	97	0.00
74 T	N-amyl acetate	1.700	1.631	4.1	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.241	1.200	3.3	95	0.00
76 T	1,2,3-Trichloropropane	1.164	1.190	-2.2	96	0.00
77 T	Bromobenzene	0.871	0.876	-0.6	97	0.00
78 T	n-propylbenzene	4.149	4.330	-4.4	99	0.00
79 T	2-Chlorotoluene	2.517	2.569	-2.1	97	0.00
80 T	1,3,5-Trimethylbenzene	3.001	3.110	-3.6	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.382	0.373	2.4	94	0.00
82 T	4-Chlorotoluene	2.988	3.064	-2.5	98	0.00
83 T	tert-Butylbenzene	2.922	2.986	-2.2	97	0.00
84 T	1,2,4-Trimethylbenzene	3.026	3.217	-6.3	100	0.00
85 T	sec-Butylbenzene	3.371	3.562	-5.7	98	0.00
86 T	p-Isopropyltoluene	3.068	3.205	-4.5	95	0.00
87 T	1,3-Dichlorobenzene	1.568	1.607	-2.5	101	0.00
88 T	1,4-Dichlorobenzene	1.622	1.615	0.4	100	0.00
89 T	n-Butylbenzene	2.811	3.004	-6.9	99	0.00
90 T	Hexachloroethane	0.554	0.563	-1.6	98	0.00
91 T	1,2-Dichlorobenzene	1.525	1.568	-2.8	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.286	0.275	3.8	91	0.00
93 T	1,2,4-Trichlorobenzene	0.999	1.027	-2.8	100	0.00
94 T	Hexachlorobutadiene	0.451	0.473	-4.9	104	0.00
95 T	Naphthalene	3.259	3.293	-1.0	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033021\
Data File : VX021618.D
Acq On : 30 Mar 2021 10:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 31 04:55:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.999	1.007	-0.8	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6