

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027800.D
 Acq On : 30 Mar 2022 11:05
 Operator : JC/MD
 Sample : VX0330WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0330WBS01

Quant Time: Mar 31 05:46:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79439	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	123174	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	109223	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57266	55.870	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 111.740%	
35) Dibromofluoromethane	5.385	113	40556	51.010	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.020%	
50) Toluene-d8	8.653	98	139092	49.722	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.440%	
62) 4-Bromofluorobenzene	11.079	95	57687	48.526	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.060%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	14888	20.683	ug/l	100
3) Chloromethane	1.294	50	11918	18.736	ug/l	100
4) Vinyl Chloride	1.374	62	12998	19.924	ug/l	96
5) Bromomethane	1.599	94	8412	28.597	ug/l	99
6) Chloroethane	1.685	64	7952	18.083	ug/l	93
7) Trichlorofluoromethane	1.886	101	21891	19.129	ug/l	96
8) Diethyl Ether	2.136	74	7567	18.421	ug/l	85
9) 1,1,2-Trichlorotrifluo...	2.331	101	15012	20.182	ug/l	97
10) Methyl Iodide	2.453	142	15191	17.586	ug/l	90
11) Tert butyl alcohol	2.959	59	17751	90.159	ug/l	99
12) 1,1-Dichloroethene	2.319	96	12367	18.120	ug/l	98
13) Acrolein	2.233	56	13422	91.021	ug/l	96
14) Allyl chloride	2.666	41	22644	18.072	ug/l	# 91
15) Acrylonitrile	3.062	53	45039	97.423	ug/l	99
16) Acetone	2.380	43	43765	96.479	ug/l	93
17) Carbon Disulfide	2.514	76	23176	16.600	ug/l	100
18) Methyl Acetate	2.703	43	19542	17.476	ug/l	92
19) Methyl tert-butyl Ether	3.111	73	50495	19.419	ug/l	98
20) Methylene Chloride	2.788	84	15351	18.814	ug/l	97
21) trans-1,2-Dichloroethene	3.093	96	13201	17.821	ug/l	94
22) Diisopropyl ether	3.757	45	50373	20.060	ug/l	# 75
23) Vinyl Acetate	3.721	43	211554	97.510	ug/l	94
24) 1,1-Dichloroethane	3.605	63	27057	19.152	ug/l	99
25) 2-Butanone	4.556	43	65217	98.806	ug/l	92
26) 2,2-Dichloropropane	4.477	77	21321	18.896	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	17110	19.155	ug/l	98
28) Bromochloromethane	4.891	49	10602	17.447	ug/l	98
29) Tetrahydrofuran	5.001	42	36792	86.776	ug/l	90
30) Chloroform	5.093	83	27955	17.801	ug/l	95
31) Cyclohexane	5.471	56	20952	18.492	ug/l	92
32) 1,1,1-Trichloroethane	5.385	97	24638	18.136	ug/l	99
36) 1,1-Dichloropropene	5.696	75	20467	20.439	ug/l	95
37) Ethyl Acetate	4.715	43	24055	19.377	ug/l	96
38) Carbon Tetrachloride	5.684	117	23148	20.923	ug/l	99
39) Methylcyclohexane	7.379	83	2			

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027800.D
 Acq On : 30 Mar 2022 11:05
 Operator : JC/MD
 Sample : VX0330WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0330WBS01

Quant Time: Mar 31 05:46:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	12040	18.234	ug/l	91
42) 1,2-Dichloroethane	6.092	62	27253	22.475	ug/l	91
43) Isopropyl Acetate	6.342	43	39711	19.972	ug/l #	90
44) Trichloroethene	7.129	130	15637	18.961	ug/l	88
45) 1,2-Dichloropropane	7.434	63	15924	19.533	ug/l	98
46) Dibromomethane	7.580	93	12378	20.174	ug/l #	87
47) Bromodichloromethane	7.824	83	23609	19.320	ug/l	98
48) Methyl methacrylate	7.690	41	20048	19.226	ug/l #	78
49) 1,4-Dioxane	7.659	88	7077	343.800	ug/l #	83
51) 4-Methyl-2-Pentanone	8.574	43	134173	97.273	ug/l	87
52) Toluene	8.720	92	36600	19.147	ug/l	97
53) t-1,3-Dichloropropene	8.982	75	22820	18.633	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	24382	18.906	ug/l #	79
55) 1,1,2-Trichloroethane	9.153	97	16354	19.290	ug/l	97
56) Ethyl methacrylate	9.116	69	25941	19.319	ug/l #	74
57) 1,3-Dichloropropane	9.311	76	27264	19.109	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.244	63	59964	89.448	ug/l	94
59) 2-Hexanone	9.433	43	102668	96.668	ug/l	82
60) Dibromochloromethane	9.525	129	16583	18.864	ug/l	97
61) 1,2-Dibromoethane	9.610	107	16869	19.042	ug/l	96
64) Tetrachloroethene	9.275	164	14813	20.002	ug/l	95
65) Chlorobenzene	10.079	112	40577	18.865	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	15360	19.031	ug/l	98
67) Ethyl Benzene	10.195	91	76109	19.500	ug/l	96
68) m/p-Xylenes	10.305	106	54181	37.743	ug/l	87
69) o-Xylene	10.640	106	26921	19.200	ug/l	86
70) Styrene	10.659	104	45105	18.926	ug/l #	92
71) Bromoform	10.805	173	10550	17.897	ug/l #	98
73) Isopropylbenzene	10.963	105	75126	19.807	ug/l	97
74) N-amyl acetate	10.841	43	33111	19.381	ug/l #	84
75) 1,1,2,2-Tetrachloroethane	11.213	83	24543	19.056	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	28011	23.440	ug/l	92
77) Bromobenzene	11.201	156	16574	19.167	ug/l	85
78) n-propylbenzene	11.305	91	88293	19.485	ug/l	97
79) 2-Chlorotoluene	11.366	91	54331	19.079	ug/l	93
80) 1,3,5-Trimethylbenzene	11.451	105	64358	19.555	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.018	75	6592	18.922	ug/l #	68
82) 4-Chlorotoluene	11.457	91	62711	19.190	ug/l	96
83) tert-Butylbenzene	11.713	119	61078	19.192	ug/l	86
84) 1,2,4-Trimethylbenzene	11.756	105	63196	19.258	ug/l	95
85) sec-Butylbenzene	11.890	105	76614	19.507	ug/l	98
86) p-Isopropyltoluene	12.012	119	63973	19.583	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	31683	18.613	ug/l	96
88) 1,4-Dichlorobenzene	12.042	146	32693	18.644	ug/l	94
89) n-Butylbenzene	12.335	91	58269	19.837	ug/l	95
90) Hexachloroethane	12.542	117	9813	18.705	ug/l	89
91) 1,2-Dichlorobenzene	12.335	146	31273	18.600	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.945	75	5934	19.139	ug/l	72
93) 1,2,4-Trichlorobenzene	13.591	180	21859	21.259	ug/l	99
94) Hexachlorobutadiene	13.725	225	10133	22.753	ug/l	94
95) Naphthalene	13.780	128	7252			

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027800.D
Acq On : 30 Mar 2022 11:05
Operator : JC/MD
Sample : VX0330WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01

Quant Time: Mar 31 05:46:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027800.D
 Acq On : 30 Mar 2022 11:05
 Operator : JC/MD
 Sample : VX0330WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0330WBS01

Quant Time: Mar 31 05:46:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion:168 Resp: 79439
Ion Ratio Lower Upper
168 100
99 62.6 41.7 62.5#



#2
Dichlorodifluoromethane
Concen: 20.683 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 85 Resp: 14888
Ion Ratio Lower Upper
85 100
87 32.4 16.2 48.6









#3
Chloromethane
Concen: 18.736 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion: 50 Resp: 11918
Ion Ratio Lower Upper
50 100
52 32.8 26.4 39.6



#4
Vinyl Chloride
Concen: 19.924 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 62 Resp: 12998
Ion Ratio Lower Upper
62 100
64 33.9 25.5 38.3









#5
Bromomethane
Concen: 28.597 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion: 94 Resp: 8412
Ion Ratio Lower Upper
94 100
96 92.8 75.3 112.9



#6
Chloroethane
Concen: 18.083 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.007 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 64 Resp: 7952
Ion Ratio Lower Upper
64 100
66 35.6 25.3 37.9









#7
 Trichlorofluoromethane
 Concen: 19.129 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX027800.D
 Acq: 30 Mar 2022 11:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0330WBS01


Tgt Ion:101 Resp: 21891
 Ion Ratio Lower Upper
 101 100
 103 66.0 50.5 75.7



#8
 Diethyl Ether
 Concen: 18.421 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX027800.D
 Acq: 30 Mar 2022 11:05

Tgt Ion: 74 Resp: 7567
 Ion Ratio Lower Upper
 74 100
 45 97.5 42.0 126.0









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.182 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion:	101	Resp:	15012
Ion	Ratio	Lower	Upper
101	100		
85	43.6	33.5	50.3
151	73.9	61.8	92.6



#10
Methyl Iodide
Concen: 17.586 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion:	142	Resp:	15191
Ion	Ratio	Lower	Upper
142	100		
127	50.4	33.9	50.9
141	15.0	11.4	17.2





#11

Tert butyl alcohol

Concen: 90.159 ug/l

RT: 2.959 min Scan# 308

Delta R.T. -0.006 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :





VX0330WBS01

Tgt Ion: 59 Resp: 17751

Ion Ratio Lower Upper

59 100

57 10.0 8.3 12.5



#12

1,1-Dichloroethene

Concen: 18.120 ug/l

RT: 2.319 min Scan# 202

Delta R.T. -0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05




Tgt Ion: 96 Resp: 12367


Ion Ratio Lower Upper

96 100

61 159.7 124.6 187.0

98 64.0 51.0 76.4









#13
Acrolein
Concen: 91.021 ug/l
RT: 2.233 min Scan# 188
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion: 56 Resp: 13422
Ion Ratio Lower Upper
56 100
55 72.0 55.0 82.6



#14
Allyl chloride
Concen: 18.072 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 41 Resp: 22644
Ion Ratio Lower Upper
41 100
39 67.5 50.4 75.6
76 32.3 32.7 49.1#





#15

Acrylonitrile

Concen: 97.423 ug/l

RT: 3.062 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :

VX0330WBS01





Tgt Ion: 53 Resp: 45039

Ion Ratio Lower Upper

53 100

52 82.4 65.7 98.5

51 37.1 28.4 42.6



#16

Acetone

Concen: 96.479 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.000 min

Lab File: VX027800.D




Acq: 30 Mar 2022 11:05


Tgt Ion: 43 Resp: 43765

Ion Ratio Lower Upper

43 100

58 31.7 28.8 43.2









#17
Carbon Disulfide
Concen: 16.600 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion: 76 Resp: 23176
Ion Ratio Lower Upper
76 100
78 8.8 7.2 10.8



#18
Methyl Acetate
Concen: 17.476 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 43 Resp: 19542
Ion Ratio Lower Upper
43 100
74 24.3 22.8 34.2









#19
Methyl tert-butyl Ether
Concen: 19.419 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion: 73 Resp: 50495
Ion Ratio Lower Upper
73 100
57 21.9 16.8 25.2



#20
Methylene Chloride
Concen: 18.814 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 84 Resp: 15351
Ion Ratio Lower Upper
84 100
49 110.5 85.5 128.3
51 34.1 26.6 39.8
86 61.1 51.5 77.3









#21
trans-1,2-Dichloroethene
Concen: 17.821 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion: 96 Resp: 13201
Ion Ratio Lower Upper
96 100
61 141.3 105.4 158.2
98 60.5 50.0 75.0



#22
Diisopropyl ether
Concen: 20.060 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 45 Resp: 50373
Ion Ratio Lower Upper
45 100
43 93.9 50.8 76.2#
87 28.8 26.6 39.8
59 13.2 10.5 15.7









#23
 Vinyl Acetate
 Concen: 97.510 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.006 min
 Lab File: VX027800.D
 Acq: 30 Mar 2022 11:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0330WBS01


Tgt Ion: 43 Resp: 211554
 Ion Ratio Lower Upper
 43 100
 86 11.0 10.8 16.2



#24
 1,1-Dichloroethane
 Concen: 19.152 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.006 min
 Lab File: VX027800.D
 Acq: 30 Mar 2022 11:05

Tgt Ion: 63 Resp: 27057
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.9 11.7
 100 4.7 2.4 7.2









#25
2-Butanone
Concen: 98.806 ug/l
RT: 4.556 min Scan# 501
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion: 43 Resp: 65217
Ion Ratio Lower Upper
43 100
72 26.3 24.4 36.6



#26
2,2-Dichloropropane
Concen: 18.896 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.006 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 77 Resp: 21321
Ion Ratio Lower Upper
77 100
97 23.2 11.8 35.4









#27
 cis-1,2-Dichloroethene
 Concen: 19.155 ug/l
 RT: 4.489 min Scan# 51
 Delta R.T. -0.006 min
 Lab File: VX027800.D
 Acq: 30 Mar 2022 11:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0330WBS01


Tgt Ion: 96 Resp: 17110
 Ion Ratio Lower Upper
 96 100
 61 144.0 0.0 283.8
 98 64.0 0.0 129.8



#28
 Bromochloromethane
 Concen: 17.447 ug/l
 RT: 4.891 min Scan# 624
 Delta R.T. -0.006 min
 Lab File: VX027800.D
 Acq: 30 Mar 2022 11:05

Tgt Ion: 49 Resp: 10602
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 4.4
 130 90.1 70.7 106.1





#29

Tetrahydrofuran

Concen: 86.776 ug/l

RT: 5.001 min Scan# 64

Delta R.T. -0.006 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :

VX0330WBS01





Tgt Ion: 42 Resp: 36792

Ion Ratio Lower Upper

42 100

72 47.5 44.1 66.1

71 43.5 40.5 60.7



#30

Chloroform

Concen: 17.801 ug/l

RT: 5.093 min Scan# 657

Delta R.T. -0.006 min

Lab File: VX027800.D




Acq: 30 Mar 2022 11:05


Tgt Ion: 83 Resp: 27955

Ion Ratio Lower Upper

83 100

85 61.8 52.7 79.1









#31
Cyclohexane
Concen: 18.492 ug/l
RT: 5.471 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion:	56	Resp:	20952
Ion Ratio	Lower	Upper	
56	100		
69	31.4	26.7	40.1
84	87.3	77.2	115.8



#32
1,1,1-Trichloroethane
Concen: 18.136 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion:	97	Resp:	24638
Ion Ratio	Lower	Upper	
97	100		
99	65.2	51.8	77.6
61	43.9	34.0	51.0





#33

1,2-Dichloroethane-d4

Concen: 55.870 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :





VX0330WBS01

Tgt Ion: 65 Resp: 57266

Ion Ratio Lower Upper

65 100

67 49.4 0.0 107.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05




Tgt Ion: 114 Resp: 123174


Ion Ratio Lower Upper

114 100

63 22.2 0.0 36.8

88 18.4 0.0 32.4





#35

Dibromofluoromethane

Concen: 51.010 ug/l

RT: 5.385 min Scan# 705

Delta R.T. -0.006 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :

VX0330WBS01





Tgt Ion:113 Resp: 40556

Ion Ratio Lower Upper

113 100

111 102.7 82.8 124.2

192 18.4 16.9 25.3



#36

1,1-Dichloropropene

Concen: 20.439 ug/l

RT: 5.696 min Scan# 756

Delta R.T. 0.006 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05




Tgt Ion: 75 Resp: 20467


Ion Ratio Lower Upper

75 100

110 32.3 18.3 54.8

77 29.8 24.8 37.2





#37

Ethyl Acetate

Concen: 19.377 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :

VX0330WBS01





Tgt Ion: 43 Resp: 24055

Ion Ratio Lower Upper

43 100

61 13.6 12.6 19.0

70 11.8 10.2 15.2



#38

Carbon Tetrachloride

Concen: 20.923 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05




Tgt Ion: 117 Resp: 23148


Ion Ratio Lower Upper

117 100

119 97.0 77.3 115.9

121 30.6 24.2 36.2





#39

Methylcyclohexane

Concen: 18.728 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. -0.006 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :

VX0330WBS01





Tgt Ion: 83 Resp: 22069

Ion Ratio Lower Upper

83 100

55 84.1 54.6 81.8#

98 49.7 37.4 56.2



#40

Benzene

Concen: 19.300 ug/l

RT: 6.044 min Scan# 813

Delta R.T. 0.007 min

Lab File: VX027800.D




Acq: 30 Mar 2022 11:05


Tgt Ion: 78 Resp: 58568

Ion Ratio Lower Upper

78 100

77 22.4 19.3 28.9









#41
Methacrylonitrile
Concen: 18.234 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion:	41	Resp:	12040
Ion	Ratio	Lower	Upper
41	100		
39	55.1	43.0	64.4
67	73.4	70.2	105.4
52	30.1	26.0	39.0



#42
1,2-Dichloroethane
Concen: 22.475 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion:	62	Resp:	27253
Ion	Ratio	Lower	Upper
62	100		
98	6.9	0.0	20.4









#43
Isopropyl Acetate
Concen: 19.972 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion: 43 Resp: 39711
Ion Ratio Lower Upper
43 100
61 18.1 18.5 27.7#
87 12.0 13.0 19.6#



#44
Trichloroethene
Concen: 18.961 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion:130 Resp: 15637
Ion Ratio Lower Upper
130 100
95 105.6 0.0 188.8





#45

1,2-Dichloropropane

Concen: 19.533 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :





VX0330WBS01

Tgt Ion: 63 Resp: 15924

Ion Ratio Lower Upper

63 100

65 33.3 25.8 38.6



#46

Dibromomethane

Concen: 20.174 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05




Tgt Ion: 93 Resp: 12378


Ion Ratio Lower Upper

93 100

95 81.3 65.8 98.6

174 82.8 85.4 128.2#









#47
Bromodichloromethane
Concen: 19.320 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion: 83	Resp: 23609		
Ion	Ratio	Lower	Upper
83	100		
85	66.7	52.2	78.2
127	8.3	6.8	10.2



#48
Methyl methacrylate
Concen: 19.226 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.006 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 41	Resp: 20048		
Ion	Ratio	Lower	Upper
41	100		
69	75.0	82.5	123.7#
39	63.3	43.8	65.8





#49

1,4-Dioxane

Concen: 343.800 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :

VX0330WBS01





Tgt Ion: 88 Resp: 7077

Ion Ratio Lower Upper

88 100

43 39.4 21.9 32.9#

58 78.2 53.7 80.5



#50

Toluene-d8

Concen: 49.722 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX027800.D




Acq: 30 Mar 2022 11:05


Tgt Ion: 98 Resp: 139092

Ion Ratio Lower Upper

98 100

100 64.7 52.6 78.8









#51
4-Methyl-2-Pentanone
Concen: 97.273 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion: 43 Resp: 134173
Ion Ratio Lower Upper
43 100
58 36.4 36.1 54.1



#52
Toluene
Concen: 19.147 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 92 Resp: 36600
Ion Ratio Lower Upper
92 100
91 175.9 137.5 206.3





#53

t-1,3-Dichloropropene

Concen: 18.633 ug/l

RT: 8.982 min Scan# 1295

Delta R.T. 0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :





VX0330WBS01

Tgt Ion: 75 Resp: 22820

Ion Ratio Lower Upper

75 100

77 31.0 25.4 38.0



#54

cis-1,3-Dichloropropene

Concen: 18.906 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05


Tgt Ion: 75 Resp: 24382


Ion Ratio Lower Upper

75 100

77 30.8 24.8 37.2

39 59.8 29.6 44.4#





#55

1,1,2-Trichloroethane

Concen: 19.290 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. -0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :

VX0330WBS01

Tgt Ion: 97 Resp: 16354





Ion Ratio Lower Upper

97 100

83 87.4 66.2 99.4

85 55.1 42.9 64.3

99 60.4 49.0 73.6



#56

Ethyl methacrylate

Concen: 19.319 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05


Tgt Ion: 69 Resp: 25941


Ion Ratio Lower Upper

69 100

41 70.6 43.8 65.6#

39 46.1 22.2 33.2#





#57

1,3-Dichloropropane

Concen: 19.109 ug/l

RT: 9.311 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :





VX0330WBS01

Tgt Ion: 76 Resp: 27264

Ion Ratio Lower Upper

76 100

78 31.0 25.8 38.8



#58

2-Chloroethyl Vinyl ether

Concen: 89.448 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX027800.D




Acq: 30 Mar 2022 11:05


Tgt Ion: 63 Resp: 59964

Ion Ratio Lower Upper

63 100

106 26.7 23.9 35.9





#59

2-Hexanone

Concen: 96.668 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :





VX0330WBS01

Tgt Ion: 43 Resp: 102668

Ion Ratio Lower Upper

43 100

58 49.5 31.9 95.7



#60

Dibromochloromethane

Concen: 18.864 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX027800.D




Acq: 30 Mar 2022 11:05


Tgt Ion: 129 Resp: 16583

Ion Ratio Lower Upper

129 100

127 78.3 40.3 120.8









#61
1,2-Dibromoethane
Concen: 19.042 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion:107 Resp: 16869
Ion Ratio Lower Upper
107 100
109 89.1 74.4 111.6



#62
4-Bromofluorobenzene
Concen: 48.526 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.006 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 95 Resp: 57687
Ion Ratio Lower Upper
95 100
174 66.3 0.0 163.0
176 64.9 0.0 155.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion:117 Resp: 109223
Ion Ratio Lower Upper
117 100
82 60.9 41.8 62.8
119 31.8 24.8 37.2



#64
Tetrachloroethene
Concen: 20.002 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion:164 Resp: 14813
Ion Ratio Lower Upper
164 100
166 123.7 101.2 151.8
129 98.4 73.0 109.4
131 99.1 74.2 111.2









#65
Chlorobenzene
Concen: 18.865 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion:112 Resp: 40577
Ion Ratio Lower Upper
112 100
114 32.1 25.8 38.8



#66
1,1,1,2-Tetrachloroethane
Concen: 19.031 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion:131 Resp: 15360
Ion Ratio Lower Upper
131 100
133 97.5 48.3 144.8
119 63.4 32.6 98.0





#67

Ethyl Benzene

Concen: 19.500 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :





VX0330WBS01

Tgt Ion: 91 Resp: 76109

Ion Ratio Lower Upper

91 100

106 28.6 24.6 36.8



#68

m/p-Xylenes

Concen: 37.743 ug/l


RT: 10.305 min Scan# 1512

Delta R.T. -0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05





Tgt Ion: 106 Resp: 54181



#69
o-Xylene
Concen: 19.200 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion:106 Resp: 26921
Ion Ratio Lower Upper
106 100
91 233.8 106.1 318.1



#70
Styrene
Concen: 18.926 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion:104 Resp: 45105
Ion Ratio Lower Upper
104 100
78 59.1 38.6 57.8#
103 55.6 43.8 65.8









#71
Bromoform
Concen: 17.897 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05


Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion:173 Resp: 10550
Ion Ratio Lower Upper
173 100
175 49.7 24.1 72.2
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion:152 Resp: 50941
Ion Ratio Lower Upper
152 100
115 75.5 37.5 112.7
150 163.5 0.0 338.2









#73
Isopropylbenzene
Concen: 19.807 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion:105 Resp: 75126
Ion Ratio Lower Upper
105 100
120 25.3 13.5 40.5



#74
N-ethyl acetate
Concen: 19.381 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 43 Resp: 33111
Ion Ratio Lower Upper
43 100
70 36.9 42.5 63.7#
55 24.2 23.1 34.7
61 22.2 22.8 34.2#









#75
 1,1,2,2-Tetrachloroethane
 Concen: 19.056 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX027800.D
 Acq: 30 Mar 2022 11:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0330WBS01


Tgt Ion: 83 Resp: 24543
 Ion Ratio Lower Upper
 83 100
 131 10.3 5.1 15.3
 85 64.5 32.6 97.7



#76
 1,2,3-Trichloropropane
 Concen: 23.440 ug/l
 RT: 11.244 min Scan# 1666
 Delta R.T. -0.000 min
 Lab File: VX027800.D
 Acq: 30 Mar 2022 11:05

Tgt Ion: 75 Resp: 28011
 Ion Ratio Lower Upper
 75 100
 77 33.7 19.4 58.2





#77

Bromobenzene

Concen: 19.167 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. 0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :

VX0330WBS01





Tgt Ion:156 Resp: 16574

Ion Ratio Lower Upper

156 100

77 186.0 78.5 235.3

158 97.1 50.2 150.7




#78

n-propylbenzene

Concen: 19.485 ug/l

RT: 11.305 min Scan# 1676



#79

2-Chlorotoluene

Concen: 19.079 ug/l

RT: 11.366 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :





VX0330WBS01


Tgt Ion: 91 Resp: 54331

Ion Ratio Lower Upper

91 100

126 30.8 17.3 51.9









#81
trans-1,4-Dichloro-2-butene
Concen: 18.922 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05




Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion: 75 Resp: 6592
Ion Ratio Lower Upper
75 100
53 129.6 69.0 103.4#
89 43.8 37.0 55.6



#82
4-Chlorotoluene
Concen: 19.190 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 91 Resp: 62711
Ion Ratio Lower Upper
91 100
126 27.3 14.9 44.5





#83

tert-Butylbenzene

Concen: 19.192 ug/l

RT: 11.713 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :

VX0330WBS01





Tgt Ion:119 Resp: 61078


Ion Ratio Lower Upper

119 100

91 67.9 26.9 80.6

134 22.2 11.4 34.2









#85
 sec-Butylbenzene
 Concen: 19.507 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX027800.D
 Acq: 30 Mar 2022 11:05


Instrument :
 MSVOA_X
 ClientSampleId :
 VX0330WBS01


Tgt Ion:105 Resp: 76614
 Ion Ratio Lower Upper
 105 100
 134 19.3 10.2 30.4



#86
 p-Isopropyltoluene
 Concen: 19.583 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX027800.D
 Acq: 30 Mar 2022 11:05

Tgt Ion:119 Resp: 63973
 Ion Ratio Lower Upper
 119 100
 134 26.1 13.0 39.0
 91 28.7 11.3 33.9





#87

1,3-Dichlorobenzene

Concen: 18.613 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :

VX0330WBS01



Tgt Ion:146 Resp: 31683


Ion Ratio Lower Upper

146 100

111 45.2 19.7 59.1

148 62.5 31.6 94.7









#89
n-Butylbenzene
Concen: 19.837 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05


Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01


Tgt Ion: 91 Resp: 58269
Ion Ratio Lower Upper
91 100
92 51.9 27.8 83.3
134 24.4 13.3 39.8



#90
Hexachloroethane
Concen: 18.705 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Tgt Ion: 117 Resp: 9813
Ion Ratio Lower Upper
117 100
201 69.4 39.5 118.5










#91
1,2-Dichlorobenzene
Concen: 18.600 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01

Tgt Ion:146 Resp: 31273
Ion Ratio Lower Upper
146 100
111 46.0 21.0 63.0
148 63.8 32.6 97.7



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.139 ug/l
RT: 1



#93

1,2,4-Trichlorobenzene

Concen: 21.259 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. -0.000 min

Lab File: VX027800.D

Acq: 30 Mar 2022 11:05

Instrument :

MSVOA_X

ClientSampleId :

VX0330WBS01



Tgt Ion:180 Resp: 21859


Ion Ratio Lower Upper

180 100

182 95.2 47.7 143.1

145 31.8 15.2 45.5





#95
Naphthalene
Concen: 20.201 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX027800.D
Acq: 30 Mar 2022 11:05

Instrument :
MSVOA_X
ClientSampleId :
VX0330WBS01

Tgt Ion:128 Resp: 72525
Ion Ratio Lower Upper
128 100
127 13.3 10.2 15.4
129 11.1 8.7 13.1

