

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027801.D
 Acq On : 30 Mar 2022 11:33
 Operator : JC/MD
 Sample : N2041-12DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D1-20220317DL

Quant Time: Mar 31 05:46:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	83961	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	143878	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50799	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59022	54.482	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.960%	
35) Dibromofluoromethane	5.391	113	47418	51.059	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.120%	
50) Toluene-d8	8.647	98	172116	52.674	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.340%	
62) 4-Bromofluorobenzene	11.079	95	63779	45.930	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 91.860%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	858	1.091	ug/l	92
12) 1,1-Dichloroethene	2.325	96	202	0.280	ug/l	# 57
27) cis-1,2-Dichloroethene	4.495	96	715	0.757	ug/l	80
44) Trichloroethene	7.129	130	42551	44.171	ug/l	94
64) Tetrachloroethene	9.275	164	1725	2.083	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027801.D
Acq On : 30 Mar 2022 11:33
Operator : JC/MD
Sample : N2041-12DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D1-20220317DL

Quant Time: Mar 31 05:46:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

