

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027804.D
 Acq On : 30 Mar 2022 12:43
 Operator : JC/MD
 Sample : N2041-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D3-20220317

Quant Time: Mar 31 05:47:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	85745	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	148098	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	129855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55416	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58307	52.702	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.400%	
35) Dibromofluoromethane	5.391	113	48771	51.019	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.040%	
50) Toluene-d8	8.653	98	169274	50.328	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.660%	
62) 4-Bromofluorobenzene	11.085	95	68182	47.702	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.400%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	1611	2.007	ug/l	95
16) Acetone	2.380	43	932	1.903	ug/l #	87
44) Trichloroethene	7.129	130	210931	212.723	ug/l	88
64) Tetrachloroethene	9.275	164	349	0.396	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027804.D
Acq On : 30 Mar 2022 12:43
Operator : JC/MD
Sample : N2041-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D3-20220317

Quant Time: Mar 31 05:47:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

