

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027805.D
 Acq On : 30 Mar 2022 13:06
 Operator : JC/MD
 Sample : N2041-25 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-20220318

Quant Time: Mar 31 05:47:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	86973	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	140773	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126157	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56973	50.769	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.540%	
35) Dibromofluoromethane	5.391	113	48654	53.545	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.100%	
50) Toluene-d8	8.653	98	166565	52.099	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.200%	
62) 4-Bromofluorobenzene	11.079	95	65435	48.162	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.320%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	450	0.553	ug/l	# 90
12) 1,1-Dichloroethene	2.319	96	394	0.527	ug/l	88
27) cis-1,2-Dichloroethene	4.489	96	495	0.506	ug/l	87
44) Trichloroethene	7.129	130	240062	254.699	ug/l	87
64) Tetrachloroethene	9.275	164	238	0.278	ug/l	# 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027805.D
Acq On : 30 Mar 2022 13:06
Operator : JC/MD
Sample : N2041-25 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-20220318

Quant Time: Mar 31 05:47:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

