

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027806.D
 Acq On : 30 Mar 2022 13:30
 Operator : JC/MD
 Sample : N2041-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP01-20220317

Quant Time: Mar 31 05:48:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	78554	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	139530	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125521	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67630	66.725	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	133.460%
35) Dibromofluoromethane	5.391	113	47999	53.295	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.580%
50) Toluene-d8	8.653	98	166342	52.493	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.980%
62) 4-Bromofluorobenzene	11.079	95	65745	48.822	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.640%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	1472	2.001	ug/l	90
16) Acetone	2.379	43	921	2.053	ug/l	96
44) Trichloroethene	7.128	130	205844	220.340	ug/l	86
64) Tetrachloroethene	9.274	164	391	0.459	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

