

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027807.D
 Acq On : 30 Mar 2022 13:53
 Operator : JC/MD
 Sample : N2041-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20220317

Quant Time: Mar 31 05:48:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	84564	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	139480	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127495	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54304	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	67699	62.046	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 124.100%	
35) Dibromofluoromethane	5.391	113	48991	54.416	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 108.840%	
50) Toluene-d8	8.653	98	168395	53.160	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.320%	
62) 4-Bromofluorobenzene	11.085	95	66604	49.477	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.960%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	1144	1.493	ug/l	88
9) 1,1,2-Trichlorotrifluo...	2.331	101	12712	16.054	ug/l	98
12) 1,1-Dichloroethene	2.319	96	4203	5.785	ug/l	93
16) Acetone	2.380	43	937	1.940	ug/l #	81
27) cis-1,2-Dichloroethene	4.495	96	2308	2.427	ug/l	89
30) Chloroform	5.099	83	2746	1.643	ug/l	92
32) 1,1,1-Trichloroethane	5.391	97	609	0.421	ug/l #	50
38) Carbon Tetrachloride	5.690	117	2614	2.087	ug/l	95
44) Trichloroethene	7.129	130	221373	237.048	ug/l	85
49) 1,4-Dioxane	7.665	88	271	11.626	ug/l #	75
55) 1,1,2-Trichloroethane	9.159	97	692	0.721	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027807.D
Acq On : 30 Mar 2022 13:53
Operator : JC/MD
Sample : N2041-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20220317

Quant Time: Mar 31 05:48:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

